

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112015\
 Data File : BE091248.D
 Acq On : 20 Nov 2015 16:22
 Operator : UM/NP
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.477

Quant Time: Nov 21 02:37:38 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 02:33:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.313	-8.7	69	0.02
3	1,4-Dioxane	0.325	0.329	-1.2	67	0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	66	0.02
5	Naphthalene	0.986	0.995	-0.9	65	0.02
6 SURR	2-Methylnaphthalene-d10	0.546	0.512	6.2	62	0.01
7	2-Methylnaphthalene	0.643	0.582	9.5	59	0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	54	0.02
9	Acenaphthylene	1.998	2.117	-6.0	56	0.00
10 C	Acenaphthene	1.368	1.459	-6.7	57	0.00
11	Fluorene	1.757	1.894	-7.8	55	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	54	0.00
13	Pentachlorophenol	0.038	0.052	-36.8#	92	0.00
14	Phenanthrene	4.619	4.532	1.9	52	0.00
15	Anthracene	4.200	4.042	3.8	53	0.00
16 SURR	Fluoranthene-d10	1.127	1.064	5.6	54	0.00
17 C	Fluoranthene	5.406	5.606	-3.7	58	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	59	0.00
19	Pyrene	4.544	4.343	4.4	56	0.00
20	Benzo(a)anthracene	1.027	0.941	8.4	56	0.00
21	Chrysene	1.183	1.241	-4.9	61	0.00
22 I	Perylene-d12	1.000	1.000	0.0	53	0.00
23	Benzo(b)fluoranthene	1.249	1.406	-12.6	58	0.00
24	Benzo(k)fluoranthene	1.281	1.437	-12.2	56	0.00
25 C	Benzo(a)pyrene	1.150	1.269	-10.3	57	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	5.087	-5.0	54	0.00
27	Dibenzo(a,h)anthracene	1.155	1.177	-1.9	53	0.01
28	Benzo(q,h,i)perylene	5.083	5.187	-2.0	53	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0