

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112015\
 Data File : BE091256.d
 Acq On : 20 Nov 2015 22:33
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.478

Quant Time: Nov 21 02:54:49 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 02:52:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.315	-9.4	70	0.00
3	1,4-Dioxane	0.325	0.344	-5.8	71	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	63	0.00
5	Naphthalene	0.986	1.016	-3.0	63	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.493	9.7	57	0.00
7	2-Methylnaphthalene	0.643	0.561	12.8	54	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	49	0.00
9	Acenaphthylene	1.998	2.104	-5.3	50	0.00
10 C	Acenaphthene	1.368	1.460	-6.7	51	0.00
11	Fluorene	1.757	1.933	-10.0	50	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	48	0.00
13	Pentachlorophenol	0.038	0.066	-73.7#	103	0.00
14	Phenanthrene	4.619	4.640	-0.5	47	0.00
15	Anthracene	4.200	4.381	-4.3	51	0.00
16 SURR	Fluoranthene-d10	1.127	1.033	8.3	46	0.00
17 C	Fluoranthene	5.406	5.754	-6.4	53	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	57	0.00
19	Pyrene	4.544	4.090	10.0	51	0.00
20	Benzo(a)anthracene	1.027	0.935	9.0	53	0.00
21	Chrysene	1.183	1.180	0.3	56	0.00
22 I	Perylene-d12	1.000	1.000	0.0	51	0.00
23	Benzo(b)fluoranthene	1.249	1.306	-4.6	52	0.00
24	Benzo(k)fluoranthene	1.281	1.413	-10.3	54	0.00
25 C	Benzo(a)pyrene	1.150	1.216	-5.7	53	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	5.070	-4.6	52	0.00
27	Dibenzo(a,h)anthracene	1.155	1.185	-2.6	52	0.00
28	Benzo(q,h,i)perylene	5.083	5.272	-3.7	52	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0