

Data Path : Z:\HPCHEM1\BNA E\Data\BE112115\
 Data File : BE091267.D
 Acq On : 21 Nov 2015 5:40
 Operator : UM/NP
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.480

Quant Time: Nov 21 06:16:19 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 04:53:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.311	-8.0	75	0.00
3	1,4-Dioxane	0.325	0.334	-2.8	74	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
5	Naphthalene	0.986	1.024	-3.9	72	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.509	6.8	67	0.00
7	2-Methylnaphthalene	0.643	0.580	9.8	64	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	61	0.00
9	Acenaphthylene	1.998	2.103	-5.3	62	0.00
10 C	Acenaphthene	1.368	1.500	-9.6	65	0.00
11	Fluorene	1.757	1.890	-7.6	61	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	58	0.00
13	Pentachlorophenol	0.038	0.059	-55.3#	113	0.00
14	Phenanthrene	4.619	4.535	1.8	56	0.00
15	Anthracene	4.200	4.123	1.8	59	0.00
16 SURR	Fluoranthene-d10	1.127	1.016	9.8	56	0.00
17 C	Fluoranthene	5.406	5.313	1.7	60	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	58	0.00
19	Pyrene	4.544	4.595	-1.1	58	0.00
20	Benzo(a)anthracene	1.027	1.019	0.8	59	0.00
21	Chrysene	1.183	1.235	-4.4	60	0.00
22 I	Perylene-d12	1.000	1.000	0.0	58	0.00
23	Benzo(b)fluoranthene	1.249	1.333	-6.7	61	0.00
24	Benzo(k)fluoranthene	1.281	1.371	-7.0	60	0.00
25 C	Benzo(a)pyrene	1.150	1.252	-8.9	62	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.869	-0.5	57	0.00
27	Dibenzo(a,h)anthracene	1.155	1.192	-3.2	60	0.00
28	Benzo(q,h,i)perylene	5.083	5.177	-1.8	59	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0