



Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE112119\
Data File : BE100746.D
Acq On : 21 Nov 2019 15:04
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 22 04:45:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE112119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 22 03:10:35 2019
Response via : Initial Calibration



TIC: BE100746.D

(22) Pentachlorophenol

16.795min (+0.008) 255.89ng

response 29607

Ion	Exp%	Act%
266.00	100	100
264.00	68.30	61.73
268.00	65.70	64.01
0.00	0.00	0.00