

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112215\
 Data File : BE091272.D
 Acq On : 21 Nov 2015 11:37
 Operator : UM/NP
 Sample : PB86691BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK91

Manual Integrations
 APPROVED

Mmdadoda
 11/23/2015 2:14:19 PM

Quant Time: Nov 22 01:18:27 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 22 00:42:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	2861	0.40	ng/ul	-0.02
4) Naphthalene-d8	9.54	136	13220	0.40	ng/ul	-0.02
8) Acenaphthene-d10	13.36	164	6812	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.08	188	15848m	0.40	ng/ul	-0.01
18) Chrysene-d12	20.23	240	18608	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	17812m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.92	96	28172	13.69	ng/uL	-0.04
6) 2-Methylnaphthalene-d10	11.06	152	5649	0.31	ng/ul	0.00
16) Fluoranthene-d10	18.07	212	17170	0.38	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112215\
Data File : BE091272.D
Acq On : 21 Nov 2015 11:37
Operator : UM/NP
Sample : PB86691BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SBLK91

Manual Integrations
APPROVED

Mmdadoda
11/23/2015 2:14:19 PM

Quant Time: Nov 22 01:18:27 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Nov 22 00:42:09 2015
Response via : Initial Calibration

