

Data Path : Z:\HPCHEM1\BNA E\Data\BE112215\
 Data File : BE091291.D
 Acq On : 22 Nov 2015 00:09
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.484

Quant Time: Nov 22 09:58:16 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 22 01:59:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.300	-4.2	69	0.00
3	1,4-Dioxane	0.325	0.343	-5.5	73	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
5	Naphthalene	0.986	1.007	-2.1	74	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.538	1.5	73	0.00
7	2-Methylnaphthalene	0.643	0.610	5.1	70	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	70	0.00
9	Acenaphthylene	1.998	2.063	-3.3	70	0.00
10 C	Acenaphthene	1.368	1.423	-4.0	71	0.00
11	Fluorene	1.757	1.706	2.9	63	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	57	0.00
13	Pentachlorophenol	0.038	0.045	-18.4	83	0.00
14	Phenanthrene	4.619	4.626	-0.2	56	0.00
15	Anthracene	4.200	4.399	-4.7	61	0.00
16 SURR	Fluoranthene-d10	1.127	1.035	8.2	56	0.00
17 C	Fluoranthene	5.406	5.346	1.1	59	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	58	0.00
19	Pyrene	4.544	4.618	-1.6	58	0.00
20	Benzo(a)anthracene	1.027	1.039	-1.2	60	0.00
21	Chrysene	1.183	1.219	-3.0	59	0.00
22 I	Perylene-d12	1.000	1.000	0.0	59	0.00
23	Benzo(b)fluoranthene	1.249	1.325	-6.1	61	0.00
24	Benzo(k)fluoranthene	1.281	1.278	0.2	56	0.00
25 C	Benzo(a)pyrene	1.150	1.226	-6.6	62	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.800	1.0	57	0.00
27	Dibenzo(a,h)anthracene	1.155	1.200	-3.9	61	0.00
28	Benzo(q,h,i)perylene	5.083	5.128	-0.9	59	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0