

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112215\
 Data File : BE091294.D
 Acq On : 22 Nov 2015 11:24
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.485

Quant Time: Nov 23 00:45:15 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 00:41:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.304	-5.6	69	0.00
3	1,4-Dioxane	0.325	0.334	-2.8	70	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
5	Naphthalene	0.986	1.014	-2.8	71	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.515	5.7	67	0.00
7	2-Methylnaphthalene	0.643	0.593	7.8	65	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	62	0.00
9	Acenaphthylene	1.998	2.121	-6.2	64	0.00
10 C	Acenaphthene	1.368	1.460	-6.7	65	0.00
11	Fluorene	1.757	1.798	-2.3	59	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	52	0.00
13	Pentachlorophenol	0.038	0.026	31.6#	45	0.01
14	Phenanthrene	4.619	4.659	-0.9	52	0.00
15	Anthracene	4.200	4.275	-1.8	55	0.00
16 SURR	Fluoranthene-d10	1.127	1.001	11.2	50	0.00
17 C	Fluoranthene	5.406	5.209	3.6	53	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	51	0.00
19	Pyrene	4.544	4.476	1.5	50	0.00
20	Benzo(a)anthracene	1.027	0.961	6.4	49	0.00
21	Chrysene	1.183	1.235	-4.4	53	0.00
22 I	Perylene-d12	1.000	1.000	0.0	49	0.00
23	Benzo(b)fluoranthene	1.249	1.312	-5.0	50	0.00
24	Benzo(k)fluoranthene	1.281	1.423	-11.1	52	0.00
25 C	Benzo(a)pyrene	1.150	1.243	-8.1	52	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.989	-2.9	49	0.00
27	Dibenzo(a,h)anthracene	1.155	1.179	-2.1	50	0.00
28	Benzo(q,h,i)perylene	5.083	5.295	-4.2	51	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0