

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112215\
 Data File : Be091280.d
 Acq On : 21 Nov 2015 17:18
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.483

Quant Time: Nov 22 01:34:39 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 22 01:30:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.306	-6.3	71	0.00
3	1,4-Dioxane	0.325	0.335	-3.1	72	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	73	0.00
5	Naphthalene	0.986	1.020	-3.4	73	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.527	3.5	70	0.00
7	2-Methylnaphthalene	0.643	0.620	3.6	69	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	67	0.00
9	Acenaphthylene	1.998	2.091	-4.7	68	0.00
10 C	Acenaphthene	1.368	1.453	-6.2	69	0.00
11	Fluorene	1.757	1.769	-0.7	62	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	57	0.00
13	Pentachlorophenol	0.038	0.054	-42.1#	100	0.00
14	Phenanthrene	4.619	4.582	0.8	55	0.00
15	Anthracene	4.200	4.316	-2.8	60	0.00
16 SURR	Fluoranthene-d10	1.127	1.007	10.6	54	0.00
17 C	Fluoranthene	5.406	5.237	3.1	57	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	56	0.00
19	Pyrene	4.544	4.666	-2.7	57	0.00
20	Benzo(a)anthracene	1.027	1.032	-0.5	58	0.00
21	Chrysene	1.183	1.228	-3.8	58	0.00
22 I	Perylene-d12	1.000	1.000	0.0	57	0.00
23	Benzo(b)fluoranthene	1.249	1.199	4.0	54	0.00
24	Benzo(k)fluoranthene	1.281	1.407	-9.8	60	0.00
25 C	Benzo(a)pyrene	1.150	1.233	-7.2	60	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.759	1.8	54	0.00
27	Dibenzo(a,h)anthracene	1.155	1.180	-2.2	58	0.00
28	Benzo(q,h,i)perylene	5.083	5.098	-0.3	57	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0