




Data Path : Z:\HPCHEM1\BNA E\DATA\BE112217\
Data File : BE094857.D
Acq On : 23 Nov 2017 6:06
Operator : SJ/JU
Sample : I6496-17
Misc :
ALS Vial : 32 Sample Multiplier: 1

Quant Time: Nov 23 23:15:49 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE111417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 23 23:11:00 2017
Response via : Initial Calibration



TIC: BE094857.D

(20) Perylene-d12 (I)

23.777min (-0.188) 0.40ng/ul

response 404193

Ion	Exp%	Act%
264.00	100	100
260.00	30.10	20.71#
265.00	103.50	21.22#
0.00	0.00	0.00