

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112217\
 Data File : BE094849.D
 Acq On : 23 Nov 2017 1:25
 Operator : SJ/JU
 Sample : I6496-09
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AC0

Manual Integrations
 APPROVED

Sohil
 11/27/2017 3:26:04 PM

Quant Time: Nov 23 03:20:22 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE111417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 23 02:50:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	672	0.40	ng/ul	-0.04
2) Naphthalene-d8	10.67	136	4721	0.40	ng/ul	-0.04
6) Acenaphthene-d10	14.50	164	3349	0.40	ng/ul	-0.02
10) Phenanthrene-d10	17.24	188	7889	0.40	ng/ul	-0.02
16) Chrysene-d12	21.42	240	8005	0.40	ng/ul	0.00
20) Perylene-d12	23.95	264	9000	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	2234	0.29	ng/ul	-0.01
14) Fluoranthene-d10	19.27	212	7319	0.27	ng/ul	-0.01
Target Compounds				Ovalue		
12) Phenanthrene	17.27	178	694	0.033	ng/ul#	85
15) Fluoranthene	19.30	202	2286	0.082	ng/ul	84
17) Pyrene	19.66	202	3140	0.125	ng/ul#	90
18) Benzo(a)anthracene	21.41	228	1898	0.078	ng/ul#	90
19) Chrysene	21.46	228	2401	0.099	ng/ul#	90
21) Benzo(b)fluoranthene	23.17	252	3244m	0.121	ng/ul	
22) Benzo(k)fluoranthene	23.22	252	977m	0.034	ng/ul	
23) Benzo(a)pyrene	23.83	252	2005	0.075	ng/ul#	91
24) Indeno(1,2,3-cd)pyrene	26.68	276	1437	0.049	ng/ul#	75
26) Benzo(g,h,i)perylene	27.52	276	1358	0.056	ng/ul#	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112217\
Data File : BE094849.D
Acq On : 23 Nov 2017 1:25
Operator : SJ/JU
Sample : I6496-09
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AC0

Manual Integrations
APPROVED

Sohil
11/27/2017 3:26:04 PM

Quant Time: Nov 23 03:20:22 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE111417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 23 02:50:08 2017
Response via : Initial Calibration

