

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112315\
 Data File : BE091304.D
 Acq On : 23 Nov 2015 17:46
 Operator : UM/NP
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.487

Quant Time: Nov 24 00:41:03 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 24 00:31:29 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	2155	0.40	ng/ul	0.00
4) Naphthalene-d8	9.56	136	10198	0.40	ng/ul	0.00
8) Acenaphthene-d10	13.37	164	4709	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.09	188	13951	0.40	ng/ul	0.00
18) Chrysene-d12	20.23	240	16144	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	13343	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.95	96	3538	2.28	ng/uL	0.00
6) 2-Methylnaphthalene-d10	11.07	152	5053	0.36	ng/ul	0.00
16) Fluoranthene-d10	18.07	212	14749	0.38	ng/ul	0.00
Target Compounds				Ovalue		
3) 1,4-Dioxane	2.97	88	3707	2.12	ng/ul#	62
5) Naphthalene	9.61	128	10416	0.41	ng/ul	99
7) 2-Methylnaphthalene	11.13	142	5826	0.36	ng/ul	87
9) Acenaphthylene	13.11	152	10218	0.43	ng/ul	99
10) Acenaphthene	13.42	153	7023	0.44	ng/ul#	86
11) Fluorene	14.42	166	9025	0.44	ng/ul	93
13) Pentachlorophenol	15.87	266	316	0.24	ng/ul	98
14) Phenanthrene	16.12	178	64676	0.40	ng/ul	99
15) Anthracene	16.21	178	59594	0.41	ng/ul	98
17) Fluoranthene	18.10	202	76543	0.41	ng/ul	88
19) Pyrene	18.49	202	69977	0.38	ng/ul#	78
20) Benzo(a)anthracene	20.20	228	15299	0.37	ng/ul	98
21) Chrysene	20.26	228	20509	0.43	ng/ul	98
23) Benzo(b)fluoranthene	21.56	252	18099	0.43	ng/ul	94
24) Benzo(k)fluoranthene	21.58	252	18716	0.44	ng/ul	93
25) Benzo(a)pyrene	22.01	252	16850	0.44	ng/ul	93
26) Indeno(1,2,3-cd)pyrene	23.71	276	65634	0.41	ng/ul#	91
27) Dibenzo(a,h)anthracene	23.66	278	15274	0.40	ng/ul#	72
28) Benzo(g,h,i)perylene	24.26	276	69584	0.41	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112315\
Data File : BE091304.D
Acq On : 23 Nov 2015 17:46
Operator : UM/NP
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.487

Quant Time: Nov 24 00:41:03 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 24 00:31:29 2015
Response via : Initial Calibration

