

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112315\
 Data File : BE091296.D
 Acq On : 23 Nov 2015 12:15
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.486

Quant Time: Nov 24 00:26:09 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 00:41:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.299	-3.8	72	0.00
3	1,4-Dioxane	0.325	0.325	0.0	73	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
5	Naphthalene	0.986	1.014	-2.8	73	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.514	5.9	69	0.00
7	2-Methylnaphthalene	0.643	0.587	8.7	66	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	65	0.00
9	Acenaphthylene	1.998	2.076	-3.9	66	0.00
10 C	Acenaphthene	1.368	1.442	-5.4	67	0.00
11	Fluorene	1.757	1.806	-2.8	62	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	55	0.00
13	Pentachlorophenol	0.038	0.040	-5.3	72	-0.01
14	Phenanthrene	4.619	4.635	-0.3	55	0.00
15	Anthracene	4.200	4.131	1.6	56	0.00
16 SURR	Fluoranthene-d10	1.127	0.991	12.1	52	0.00
17 C	Fluoranthene	5.406	5.132	5.1	55	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
19	Pyrene	4.544	4.543	0.0	51	0.00
20	Benzo(a)anthracene	1.027	0.957	6.8	50	0.00
21	Chrysene	1.183	1.228	-3.8	53	0.00
22 I	Perylene-d12	1.000	1.000	0.0	49	0.00
23	Benzo(b)fluoranthene	1.249	1.382	-10.6	53	0.00
24	Benzo(k)fluoranthene	1.281	1.361	-6.2	50	0.00
25 C	Benzo(a)pyrene	1.150	1.250	-8.7	53	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.988	-2.9	49	0.00
27	Dibenzo(a,h)anthracene	1.155	1.146	0.8	49	0.00
28	Benzo(q,h,i)perylene	5.083	5.216	-2.6	50	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0