

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112315\
 Data File : BE091304.D
 Acq On : 23 Nov 2015 17:46
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.487

Quant Time: Nov 24 00:41:03 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 24 00:31:29 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	54	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.328	-13.9	58	0.00
3	1,4-Dioxane	0.325	0.344	-5.8	56	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	54	0.00
5	Naphthalene	0.986	1.021	-3.5	55	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.495	9.3	49	0.00
7	2-Methylnaphthalene	0.643	0.571	11.2	48	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	47	0.00
9	Acenaphthylene	1.998	2.170	-8.6	50	0.00
10 C	Acenaphthene	1.368	1.491	-9.0	50	0.00
11	Fluorene	1.757	1.917	-9.1	48	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	46	0.00
13	Pentachlorophenol	0.038	0.023	39.5#	34	0.00
14	Phenanthrene	4.619	4.636	-0.4	46	0.00
15	Anthracene	4.200	4.272	-1.7	49	0.00
16 SURR	Fluoranthene-d10	1.127	1.057	6.2	46	0.00
17 C	Fluoranthene	5.406	5.487	-1.5	49	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	49	0.00
19	Pyrene	4.544	4.335	4.6	47	0.00
20	Benzo(a)anthracene	1.027	0.948	7.7	47	0.00
21	Chrysene	1.183	1.270	-7.4	52	0.00
22 I	Perylene-d12	1.000	1.000	0.0	48	0.00
23	Benzo(b)fluoranthene	1.249	1.356	-8.6	51	0.00
24	Benzo(k)fluoranthene	1.281	1.403	-9.5	50	0.00
25 C	Benzo(a)pyrene	1.150	1.263	-9.8	52	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.919	-1.5	47	0.00
27	Dibenzo(a,h)anthracene	1.155	1.145	0.9	47	0.00
28	Benzo(q,h,i)perylene	5.083	5.215	-2.6	49	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0