

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112315\
 Data File : BE091313.D
 Acq On : 23 Nov 2015 23:15
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.488

Quant Time: Nov 24 02:55:30 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 24 02:51:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.305	-5.9	60	0.00
3	1,4-Dioxane	0.325	0.343	-5.5	63	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
5	Naphthalene	0.986	1.033	-4.8	67	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.542	0.7	65	0.00
7	2-Methylnaphthalene	0.643	0.627	2.5	64	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	64	0.00
9	Acenaphthylene	1.998	2.168	-8.5	67	0.00
10 C	Acenaphthene	1.368	1.504	-9.9	69	0.00
11	Fluorene	1.757	1.871	-6.5	63	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	58	0.00
13	Pentachlorophenol	0.038	0.040	-5.3	75	-0.01
14	Phenanthrene	4.619	4.791	-3.7	59	0.00
15	Anthracene	4.200	4.461	-6.2	63	-0.01
16 SURR	Fluoranthene-d10	1.127	1.073	4.8	59	0.00
17 C	Fluoranthene	5.406	5.539	-2.5	62	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
19	Pyrene	4.544	4.603	-1.3	61	0.00
20	Benzo(a)anthracene	1.027	1.075	-4.7	65	0.00
21	Chrysene	1.183	1.250	-5.7	64	0.00
22 I	Perylene-d12	1.000	1.000	0.0	67	0.00
23	Benzo(b)fluoranthene	1.249	1.210	3.1	63	0.00
24	Benzo(k)fluoranthene	1.281	1.360	-6.2	68	0.00
25 C	Benzo(a)pyrene	1.150	1.234	-7.3	70	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.791	1.2	64	0.00
27	Dibenzo(a,h)anthracene	1.155	1.175	-1.7	67	0.00
28	Benzo(q,h,i)perylene	5.083	5.050	0.6	66	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0