

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112315\
 Data File : BE091313.D
 Acq On : 23 Nov 2015 23:15
 Operator : UM/NP
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.488

Quant Time: Nov 24 02:55:30 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 24 02:51:08 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	2415	0.40	ng/ul	0.00
4) Naphthalene-d8	9.56	136	12362	0.40	ng/ul	0.00
8) Acenaphthene-d10	13.36	164	6379	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.09	188	17373	0.40	ng/ul	0.00
18) Chrysene-d12	20.23	240	19939	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	18506	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.95	96	3683	2.12	ng/uL	0.00
6) 2-Methylnaphthalene-d10	11.06	152	6700	0.40	ng/ul	0.00
16) Fluoranthene-d10	18.07	212	18648	0.38	ng/ul	0.00
Target Compounds				Ovalue		
3) 1,4-Dioxane	2.97	88	4147	2.11	ng/ul#	65
5) Naphthalene	9.60	128	12772	0.42	ng/ul	100
7) 2-Methylnaphthalene	11.13	142	7757	0.39	ng/ul	88
9) Acenaphthylene	13.12	152	13827	0.43	ng/ul	98
10) Acenaphthene	13.42	153	9591	0.44	ng/ul	86
11) Fluorene	14.41	166	11933	0.43	ng/ul	97
13) Pentachlorophenol	15.86	266	690	0.42	ng/ul	95
14) Phenanthrene	16.13	178	83234	0.41	ng/ul	97
15) Anthracene	16.20	178	77500	0.42	ng/ul	97
17) Fluoranthene	18.09	202	96227	0.41	ng/ul	90
19) Pyrene	18.48	202	91780	0.41	ng/ul#	80
20) Benzo(a)anthracene	20.20	228	21435	0.42	ng/ul	97
21) Chrysene	20.26	228	24917	0.42	ng/ul	99
23) Benzo(b)fluoranthene	21.55	252	22390	0.39	ng/ul	94
24) Benzo(k)fluoranthene	21.58	252	25161	0.42	ng/ul	94
25) Benzo(a)pyrene	22.01	252	22828	0.43	ng/ul	92
26) Indeno(1,2,3-cd)pyrene	23.70	276	88655	0.40	ng/ul#	90
27) Dibenzo(a,h)anthracene	23.66	278	21737	0.41	ng/ul#	71
28) Benzo(g,h,i)perylene	24.25	276	93447	0.40	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112315\
Data File : BE091313.D
Acq On : 23 Nov 2015 23:15
Operator : UM/NP
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.488

Quant Time: Nov 24 02:55:30 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 24 02:51:08 2015
Response via : Initial Calibration

