

Data Path : Z:\HPCHEM1\BNA_E\Data\BE112316\
 Data File : BE092577.D
 Acq On : 23 Nov 2016 13:33
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 23 16:07:51 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	96	-0.02
2	1,4-Dioxane	0.400	0.475	-18.7	125	0.00
3	n-Nitrosodimethylamine	0.400	0.451	-12.7	102	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	88	0.00
5 S	Phenol-d6	0.400	0.395	1.3	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.329	17.8	90	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.390	2.5	93	0.00
9	Naphthalene	0.400	0.403	-0.8	97	-0.02
10	Hexachlorobutadiene	0.400	0.406	-1.5	96	0.00
11	2-Methylnaphthalene	0.400	0.387	3.3	98	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.400	0.365	8.8	101	0.00
14 S	2-Fluorobiphenyl	0.400	0.380	5.0	98	0.00
15	Acenaphthylene	0.400	0.371	7.3	101	0.00
16	Acenaphthene	0.400	0.387	3.3	101	0.00
17	Fluorene	0.400	0.383	4.3	100	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	97	0.00
19	Hexachlorobenzene	0.400	0.407	-1.7	111	0.00
20	Pentachlorophenol	0.400	0.433	-8.2	113	0.00
21	Phenanthrene	0.400	0.400	0.0	101	-0.02
22	Anthracene	0.400	0.382	4.5	99	0.00
23	Fluoranthene	0.400	0.391	2.3	102	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	99	0.00
25	Pyrene	0.400	0.380	5.0	100	0.00
26 S	Terphenyl-d14	0.400	0.368	8.0	98	0.00
27	Benzo(a)anthracene	0.400	0.459	-14.8	108	0.00
28	Chrysene	0.400	0.422	-5.5	100	-0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.438	-9.5	102	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	89	0.00
31 I	Perylene-d12	5.000	5.000	0.0	93	-0.01
32	Benzo(b)fluoranthene	0.400	0.356	11.0	86	-0.01
33	Benzo(k)fluoranthene	0.400	0.396	1.0	97	-0.01
34 C	Benzo(a)pyrene	0.400	0.366	8.5	92	0.00
35	Dibenzo(a,h)anthracene	0.400	0.355	11.3	89	-0.02
36	Benzo(g,h,i)perylene	0.400	0.370	7.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0