


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE112316\
 Data File : BE092571.D
 Acq On : 23 Nov 2016 9:07
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111616.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 16 15:43:18 2016



AutoFind: Scans 2532, 2533, 2534; Background Corrected with Scan 2524

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.6	20197	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	47.8	16761	PASS
70	69	0.00	2	0.4	59	PASS
127	198	10	80	53.5	18749	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	35047	PASS
199	198	5	9	5.6	1972	PASS
275	198	10	60	21.6	7566	PASS
365	198	1	100	2.7	941	PASS
441	442	0.01	24	17.4	3570	PASS
442	198	50	100	58.6	20541	PASS
443	442	15	24	17.9	3676	PASS