

Data Path : Z:\HPCHEM1\BNA_E\Data\BE112415\
 Data File : BE091321.D
 Acq On : 24 Nov 2015 12:31
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.491

Quant Time: Nov 24 13:17:41 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 24 02:51:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.316	-9.7	60	0.00
3	1,4-Dioxane	0.325	0.350	-7.7	62	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	60	0.00
5	Naphthalene	0.986	1.022	-3.7	60	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.502	8.1	55	0.00
7	2-Methylnaphthalene	0.643	0.581	9.6	53	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	52	0.00
9	Acenaphthylene	1.998	2.181	-9.2	55	0.00
10 C	Acenaphthene	1.368	1.560	-14.0	58	0.00
11	Fluorene	1.757	2.032	-15.7	56	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	52	0.00
13	Pentachlorophenol	0.038	0.019	50.0#	32	0.00
14	Phenanthrene	4.619	4.674	-1.2	52	0.00
15	Anthracene	4.200	4.165	0.8	53	0.00
16 SURR	Fluoranthene-d10	1.127	0.980	13.0	48	0.00
17 C	Fluoranthene	5.406	5.156	4.6	51	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	49	0.00
19	Pyrene	4.544	4.420	2.7	47	0.00
20	Benzo(a)anthracene	1.027	0.883	14.0	43	0.00
21	Chrysene	1.183	1.278	-8.0	52	0.00
22 I	Perylene-d12	1.000	1.000	0.0	45	0.00
23	Benzo(b)fluoranthene	1.249	1.348	-7.9	48	0.00
24	Benzo(k)fluoranthene	1.281	1.446	-12.9	49	0.00
25 C	Benzo(a)pyrene	1.150	1.244	-8.2	48	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	5.035	-3.9	45	0.00
27	Dibenzo(a,h)anthracene	1.155	1.112	3.7	43	0.00
28	Benzo(g,h,i)perylene	5.083	5.145	-1.2	45	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0