

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE112717\
 Data File : BE094864.D
 Acq On : 27 Nov 2017 10:05
 Operator : SJ/JU
 Sample : I6336-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 50

Instrument :
 BNA_E
 ClientSampleId :
 CF-1B(1-2)

Manual Integrations
 APPROVED

Sohil
 11/27/2017 5:16:13 PM

Quant Time: Nov 27 12:34:04 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	30020	20.00	ng	-0.02
7) Naphthalene-d8	10.67	136	187864	20.00	ng	-0.03
13) Acenaphthene-d10	14.50	164	123759	20.00	ng	-0.01
19) Phenanthrene-d10	17.23	188	334406	20.00	ng	0.00
27) Chrysene-d12	21.41	240	371792	20.00	ng	-0.01
34) Perylene-d12	23.94	264	360804	20.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.48	112	131148	63.76	ng	-0.05
5) Phenol-d6	7.08	99	139398	44.90	ng	-0.06
8) Nitrobenzene-d5	9.03	82	309235	103.09	ng	-0.05
11) 2-Methylnaphthalene-d10	12.22	152	1	0.00	ng	-0.07
14) 2,4,6-Tribromophenol	16.00	330	161525	236.58	ng	-0.03
15) 2-Fluorobiphenyl	13.12	172	778606	89.25	ng	-0.03
25) Fluoranthene-d10	19.29	212	92	-50.00	ng	0.00
29) Terphenyl-d14	19.84	244	1330989	94.93	ng	-0.02
Target Compounds						
9) Naphthalene	10.72	128	35295	0.830	ng	Qvalue 100
12) 2-Methylnaphthalene	12.36	142	446	0.062	ng	# 93
16) Acenaphthylene	14.23	152	1357	0.104	ng	# 94
17) Acenaphthene	14.55	154	897	0.107	ng	98
18) Fluorene	15.56	166	787	0.073	ng	96
24) Anthracene	17.40	178	483	0.025	ng	95
28) Pyrene	19.67	202	2133	0.082	ng	96
30) Benzo(a)anthracene	21.40	228	1831	0.076	ng	# 67
32) Bis(2-ethylhexyl)phthalate	21.27	149	4769	0.080	ng	# 100
33) Indeno(1,2,3-cd)pyrene	26.72	276	477	0.021	ng	# 86
35) Benzo(b)fluoranthene	23.18	252	881	0.038	ng	# 73
37) Benzo(a)pyrene	23.83	252	584m	0.026	ng	
39) Benzo(g,h,i)perylene	27.56	276	632	0.033	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE112717\
Data File : BE094864.D
Acq On : 27 Nov 2017 10:05
Operator : SJ/JU
Sample : I6336-01
Misc :
ALS Vial : 3 Sample Multiplier: 50

Instrument :
BNA_E
ClientSampleId :
CF-1B(1-2)

Manual Integrations
APPROVED

Sohil
11/27/2017 5:16:13 PM

Quant Time: Nov 27 12:34:04 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

