

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112817\
 Data File : BE094878.D
 Acq On : 28 Nov 2017 11:39
 Operator : SJ/JU
 Sample : PB104369BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK69

Quant Time: Nov 29 01:13:02 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE111417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 29 01:10:42 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	484	0.40	ng/ul	-0.05
2) Naphthalene-d8	10.68	136	3507	0.40	ng/ul	-0.04
6) Acenaphthene-d10	14.50	164	2510	0.40	ng/ul	-0.02
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.26
16) Chrysene-d12	21.45	240	5921	0.40	ng/ul	0.02
20) Perylene-d12	23.97	264	7920	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.32	152	1814	0.32	ng/ul	0.02
14) Fluoranthene-d10	19.27	212	6603	0.00	ng/ul	-0.01
Target Compounds						
9) Fluorene	15.49	166	575	0.06	ng/ul#	25
17) Pyrene	19.63	202	773	0.04	ng/ul#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112817\
Data File : BE094878.D
Acq On : 28 Nov 2017 11:39
Operator : SJ/JU
Sample : PB104369BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SBLK69

Quant Time: Nov 29 01:13:02 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE111417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 29 01:10:42 2017
Response via : Initial Calibration





#9

Fluorene

Concen: 0.06 ng/ul

RT: 15.49 min Scan# 2289

Delta R.T. -0.08 min

Lab File: BE094878.D

Acq: 28 Nov 2017 11:39

Instrument :

BNA_E

ClientSampleId :

SBLK69


Tot Ion:166 Resp: 575

Ion Ratio Lower Upper

166 100

165 14.0 73.4 110.2#




167 36.9 14.6 22.0#



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#17

Pyrene

Concen: 0.04 ng/ul

RT: 19.63 min Scan# 2822

Delta R.T. -0.06 min

Lab File: BE094878.D

Acq: 28 Nov 2017 11:39


Tgt Ion:202 Resp: 773

Ion Ratio Lower Upper

202 100

101 141.9 18.2 27.4#

100 763.4 15.6 23.4#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): E

