

Data Path : Z:\HPCHEM1\BNA E\Data\BE113017\
 Data File : BE094892.D
 Acq On : 30 Nov 2017 15:00
 Operator : SJ/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6

Quant Time: Nov 30 15:48:31 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 30 15:30:59 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.484	152	8346	0.40	ng	0.00
7) Naphthalene-d8	11.326	136	18244	0.40	ng	0.00
13) Acenaphthene-d10	15.076	164	10507	0.40	ng	0.00
19) Phenanthrene-d10	17.774	188	25079	0.40	ng	0.00
27) Chrysene-d12	21.886	240	30612	0.40	ng	0.00
34) Perylene-d12	24.567	264	23617	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.953	112	21441	0.75	ng	0.00
5) Phenol-d6	7.590	99	32333	0.75	ng	0.00
8) Nitrobenzene-d5	9.665	82	19372	1.33	ng	0.00
11) 2-Methylnaphthalene-d10	12.871	152	49066	1.68	ng	0.00
14) 2,4,6-Tribromophenol	16.533	330	3945	1.36	ng	0.00
15) 2-Fluorobiphenyl	13.713	172	73039	1.97	ng	0.00
25) Fluoranthene-d10	19.759	212	572852	1.95	ng	0.00
29) Terphenyl-d14	20.329	244	94030	1.63	ng	0.00


Target Compounds						Qvalue
2) 1,4-Dioxane	3.675	88	13574	0.96	ng	# 86
3) n-Nitrosodimethylamine	4.015	42	19070	1.41	ng	# 78
6) bis(2-Chloroethyl)ether	7.880	93	31448	0.94	ng	99
9) Naphthalene	11.377	128	353523	1.71	ng	99
10) Hexachlorobutadiene	11.644	225	14881	1.99	ng	96
12) 2-Methylnaphthalene	12.937	142	85824	2.45	ng	96
16) Acenaphthylene	14.794	152	99407	1.79	ng	99
17) Acenaphthene	15.135	154	63339	1.78	ng	97
18) Fluorene	16.099	166	80478	1.76	ng	# 78
20) 4-Bromophenyl-phenylether	16.967	248	23811	1.74	ng	# 64
21) Hexachlorobenzene	17.086	284	23124	1.93	ng	# 78
22) Pentachlorophenol	17.430	266	6096	3.98	ng	# 70
23) Phenanthrene	17.819	178	133446	1.86	ng	98
24) Anthracene	17.909	178	139757	1.90	ng	# 95
26) Fluoranthene	19.789	202	175350	1.97	ng	100
28) Pyrene	20.143	202	177440	1.66	ng	99
30) Benzo(a)anthracene	21.870	228	155579	1.56	ng	# 61
31) Chrysene	21.917	228	171198	1.69	ng	# 64
32) Bis(2-ethylhexyl)phtha...	21.761	149	266621	1.09	ng	# 100
33) Indeno(1,2,3-cd)pyrene	27.448	276	143432	1.53	ng	99
35) Benzo(b)fluoranthene	23.734	252	142262	1.86	ng	# 39
36) Benzo(k)fluoranthene	23.793	252	149206	1.83	ng	# 39
37) Benzo(a)pyrene	24.444	252	128881	1.74	ng	# 32
38) Dibenzo(a,h)anthracene	27.475	278	122111	1.83	ng	# 42
39) Benzo(g,h,i)perylene	28.336	276	118389	1.90	ng	# 52


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\Data\BE113017\
Data File : BE094892.D
Acq On : 30 Nov 2017 15:00
Operator : SJ/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6

Quant Time: Nov 30 15:48:31 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 30 15:30:59 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.40 ng

RT: 8.484 min Scan# 44

Delta R.T. -0.003 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:152 Resp: 8346

Ion Ratio Lower Upper

152 100

150 147.5 117.2 175.8

115 67.5 57.0 85.6



#2

1,4-Dioxane

Concen: 0.96 ng

RT: 3.675 min Scan# 65

Delta R.T. -0.003 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion: 88 Resp: 13574




Ion Ratio Lower Upper

88 100

58 86.4 63.2 94.8

43 35.6 41.2 61.8#





#3

n-Nitrosodimethylamine

Concen: 1.41 ng

RT: 4.015 min Scan# 92

Delta R.T. -0.015 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion: 42 Resp: 19070

Ion Ratio Lower Upper

42 100

74 102.1 100.3 150.5

44 7.1 17.0 25.4#



#4

2-Fluorophenol

Concen: 0.75 ng

RT: 5.953 min Scan# 246

Delta R.T. -0.003 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00


Tgt Ion: 112 Resp: 21441


Ion Ratio Lower Upper

112 100

64 78.1 60.1 90.1

63 162.0 116.8 175.2





#5

Phenol-d6

Concen: 0.75 ng

RT: 7.590 min Scan# 37

Delta R.T. -0.003 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion: 99 Resp: 32333

Ion Ratio Lower Upper

99 100

42 26.5 20.2 30.2

71 39.3 28.4 42.6



#6

bis(2-Chloroethyl)ether

Concen: 0.94 ng

RT: 7.880 min Scan# 399

Delta R.T. -0.003 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion: 93 Resp: 31448


Ion Ratio Lower Upper

93 100

63 346.3 275.3 412.9

95 31.8 25.2 37.8





#7

Naphthalene-d8

Concen: 0.40 ng

RT: 11.326 min Scan# 61

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6

Tot Ion:136 Resp: 18244





Ion Ratio Lower Upper

136 100

137 11.3 9.5 14.3

54 50.3 29.1 43.7#

68 9.9 6.2 9.2#



#8

Nitrobenzene-d5

Concen: 1.33 ng

RT: 9.665 min Scan# 514

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion: 82 Resp: 19372


Ion Ratio Lower Upper

82 100

128 129.6 150.0 225.0#

54 194.3 154.2 231.4





#9

Naphthalene

Concen: 1.71 ng

RT: 11.377 min Scan# 61

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:128 Resp: 353523

Ion Ratio Lower Upper

128 100

129 2.6 2.6 3.8

127 3.3 2.8 4.2



#10

Hexachlorobutadiene

Concen: 1.99 ng

RT: 11.644 min Scan# 643

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:225 Resp: 14881


Ion Ratio Lower Upper

225 100

223 62.4 53.0 79.6

227 62.7 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 1.68 ng
RT: 12.871 min Scan# 846
Delta R.T. 0.000 min
Lab File: BE094892.D
Acq: 30 Nov 2017 15:00




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6


Tot Ion:152 Resp: 49066
Ion Ratio Lower Upper
152 100
151 19.5 15.4 23.0



#12
2-Methylnaphthalene
Concen: 2.45 ng
RT: 12.937 min Scan# 857
Delta R.T. 0.000 min
Lab File: BE094892.D
Acq: 30 Nov 2017 15:00

Tgt Ion:142 Resp: 85824
Ion Ratio Lower Upper
142 100
141 92.2 70.4 105.6
115 41.0 31.7 47.5





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 15.076 min Scan# 10

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:164 Resp: 10507

Ion Ratio Lower Upper

164 100

162 102.8 83.1 124.7

160 42.6 37.1 55.7



#14

2,4,6-Tribromophenol

Concen: 1.36 ng

RT: 16.533 min Scan# 1100

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:330 Resp: 3945




Ion Ratio Lower Upper

330 100

332 98.5 152.7 229.1#

141 54.8 33.7 50.5#





#15

2-Fluorobiphenyl

Concen: 1.97 ng

RT: 13.713 min Scan# 91

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :





BNA_E

ClientSampleId :

SSTDICC1.6

Tot Ion:172 Resp: 73039

Ion	Ratio	Lower	Upper
172	100		
171	34.9	29.9	44.9
170	22.0	21.8	32.8



#16

Acenaphthylene

Concen: 1.79 ng

RT: 14.794 min Scan# 986


Delta R.T. 0.000 min


Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Tgt Ion:152 Resp: 99407

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.7	23.5
153	12.7	10.5	15.7





#17

Acenaphthene

Concen: 1.78 ng

RT: 15.135 min Scan# 1009

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:154 Resp: 63339

Ion Ratio Lower Upper

154 100

153 113.1 89.2 133.8

152 56.5 42.0 63.0



#18

Fluorene

Concen: 1.76 ng

RT: 16.099 min Scan# 1071

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:166 Resp: 80478


Ion Ratio Lower Upper

166 100

165 100.7 77.8 116.6

167 13.2 42.4 63.6#





#19

Phenanthrene-d10

Concen: 0.40 ng

RT: 17.774 min Scan# 11

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:188 Resp: 25079

Ion Ratio Lower Upper

188 100

94 0.0 3.9 5.9#

80 17.9 3.6 5.4#



#20

4-Bromophenyl-phenylether

Concen: 1.74 ng

RT: 16.967 min Scan# 1129

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:248 Resp: 23811


Ion Ratio Lower Upper

248 100

250 87.1 62.7 94.1

141 112.4 48.2 72.2#





#21

Hexachlorobenzene

Concen: 1.93 ng

RT: 17.086 min Scan# 1137

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:284 Resp: 23124

Ion Ratio Lower Upper

284 100

142 47.0 22.1 33.1#

249 35.2 34.8 52.2



#22

Pentachlorophenol

Concen: 3.98 ng

RT: 17.430 min Scan# 1160

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:266 Resp: 6096


Ion Ratio Lower Upper

266 100

264 60.4 72.5 108.7#

268 64.6 73.8 110.6#





#23

Phenanthrene

Concen: 1.86 ng

RT: 17.819 min Scan# 1186

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:178 Resp: 133446

Ion Ratio Lower Upper

178 100

176 19.7 15.1 22.7

179 15.3 11.8 17.6



#24

Anthracene

Concen: 1.90 ng

RT: 17.909 min Scan# 1192

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:178 Resp: 139757


Ion Ratio Lower Upper

178 100

176 19.5 12.8 19.2#

179 15.7 13.0 19.6









#25
Fluoranthene-d10
Concen: 1.95 ng
RT: 19.759 min Scan# 1432
Delta R.T. 0.001 min
Lab File: BE094892.D
Acq: 30 Nov 2017 15:00




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6


Tot Ion	Ratio	Lower	Upper
212	100		
106	3.6	2.8	4.2
104	2.0	1.6	2.4



#26
Fluoranthene
Concen: 1.97 ng
RT: 19.789 min Scan# 1436
Delta R.T. 0.001 min
Lab File: BE094892.D
Acq: 30 Nov 2017 15:00

Tgt Ion	Ratio	Lower	Upper
202	100		
101	12.4	9.8	14.8
203	17.2	13.7	20.5





#27

Chrysene-d12

Concen: 0.40 ng

RT: 21.886 min Scan# 1583

Delta R.T. 0.001 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:240 Resp: 30612

Ion Ratio Lower Upper

240 100

120 6.1 5.5 8.3

236 26.9 20.7 31.1



#28

Pyrene

Concen: 1.66 ng

RT: 20.143 min Scan# 1471

Delta R.T. 0.001 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:202 Resp: 177440


Ion Ratio Lower Upper

202 100

200 21.6 16.6 25.0

203 17.6 13.8 20.8





#29

Terphenyl-d14

Concen: 1.63 ng

RT: 20.329 min Scan# 14

Delta R.T. 0.001 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:244 Resp: 94030

Ion Ratio Lower Upper

244 100

212 7.2 25.4 38.0#

122 3.9 8.1 12.1#



#30

Benzo(a)anthracene

Concen: 1.56 ng

RT: 21.870 min Scan# 1582

Delta R.T. 0.001 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:228 Resp: 155579


Ion Ratio Lower Upper

228 100

226 26.2 40.0 60.0#

229 20.5 40.0 60.0#





#31

Chrysene

Concen: 1.69 ng

RT: 21.917 min Scan# 15

Delta R.T. 0.001 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:228 Resp: 171198

Ion Ratio Lower Upper

228 100

226 31.9 40.0 60.0#

229 18.8 40.0 60.0#



#32

Bis(2-ethylhexyl)phthalate

Concen: 1.09 ng

RT: 21.761 min Scan# 1575

Delta R.T. 0.001 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:149 Resp: 266621


Ion Ratio Lower Upper

149 100

167 6.6 5.4 8.0

279 0.7 0.7 1.1#









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 1.53 ng
 RT: 27.448 min Scan# 2774
 Delta R.T. 0.012 min
 Lab File: BE094892.D
 Acq: 30 Nov 2017 15:00




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC1.6


Tot Ion	Ratio	Lower	Upper
276	100		
138	29.1	22.5	33.7
277	24.8	19.9	29.9



#34
 Perylene-d12
 Concen: 0.40 ng
 RT: 24.567 min Scan# 2148
 Delta R.T. 0.000 min
 Lab File: BE094892.D
 Acq: 30 Nov 2017 15:00

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.6	20.5	30.7
265	32.9	22.8	34.2





#35

Benzo(b)fluoranthene

Concen: 1.86 ng

RT: 23.734 min Scan# 19

Delta R.T. 0.000 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6





Tot Ion:252 Resp: 142262

Ion Ratio Lower Upper

252 100

253 21.8 48.0 72.0#

125 12.0 56.0 84.0#



#36

Benzo(k)fluoranthene

Concen: 1.83 ng

RT: 23.793 min Scan# 1978

Delta R.T. 0.005 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00




Tgt Ion:252 Resp: 149206


Ion Ratio Lower Upper

252 100

253 22.2 48.0 72.0#

125 11.9 56.0 84.0#









#37
Benzo(a)pyrene
Concen: 1.74 ng
RT: 24.444 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BE094892.D
Acq: 30 Nov 2017 15:00




Instrument :
BNA_E
ClientSampleId :
SSTDICC1.6




Tot Ion:252 Resp: 128881
Ion Ratio Lower Upper
252 100
253 21.8 52.0 78.0#
125 13.1 68.0 102.0#



#38
Dibenzo(a,h)anthracene
Concen: 1.83 ng
RT: 27.475 min Scan# 2783
Delta R.T. 0.003 min
Lab File: BE094892.D
Acq: 30 Nov 2017 15:00

Tgt Ion:278 Resp: 122111
Ion Ratio Lower Upper
278 100
139 18.3 56.0 84.0#
279 23.5 52.0 78.0#





#39

Benzo(a,h,i)perylene

Concen: 1.90 ng

RT: 28.336 min Scan# 29

Delta R.T. 0.008 min

Lab File: BE094892.D

Acq: 30 Nov 2017 15:00

Instrument :

BNA_E

ClientSampleId :

SSTDICC1.6

Tot Ion: 276 Resp: 118389

Ion Ratio Lower Upper

276 100

277 23.5 52.0 78.0#

138 24.8 44.0 66.0#

