

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120115\
 Data File : BE091328.D
 Acq On : 1 Dec 2015 15:23
 Operator : UM/NP
 Sample : SSTD3.297
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD3.297

Quant Time: Dec 01 16:02:55 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE120115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 01 15:48:03 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	2607	0.40	ng/ul	-0.03
4) Naphthalene-d8	9.55	136	12917	0.40	ng/ul	-0.02
8) Acenaphthene-d10	13.36	164	5346	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.09	188	15500	0.40	ng/ul	0.00
18) Chrysene-d12	20.22	240	19735	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	11730	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
6) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
16) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
13) Pentachlorophenol	15.84	266	2436	3.72	ng/ul	Ovalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120115\
Data File : BE091328.D
Acq On : 1 Dec 2015 15:23
Operator : UM/NP
Sample : SST3.297
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD3.297

Quant Time: Dec 01 16:02:55 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE120115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 01 15:48:03 2015
Response via : Initial Calibration

