

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE120116\
 Data File : BE092587.D
 Acq On : 1 Dec 2016 14:36
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Dec 02 01:01:18 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	8138	5.00	ng	-0.02
7) Naphthalene-d8	10.93	136	41115	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	30160	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	66559	5.00	ng	0.00
24) Chrysene-d12	21.72	240	95773	5.00	ng	0.00
31) Perylene-d12	24.08	264	80993	5.00	ng	-0.01

System Monitoring Compounds						
4) 2-Fluorophenol	5.68	112	474	0.34	ng	0.00
5) Phenol-d6	7.36	99	614	0.38	ng	0.00
8) Nitrobenzene-d5	9.36	82	723	0.35	ng	0.00
13) 2,4,6-Tribromophenol	16.31	330	203	0.30	ng	0.00
14) 2-Fluorobiphenyl	13.43	172	2410	0.33	ng	0.00
26) Terphenyl-d14	20.17	244	3501	0.32	ng	0.00


Target Compounds				Qvalue		
2) 1,4-Dioxane	3.39	88	498	0.41	ng	89
3) n-Nitrosodimethylamine	3.79	42	332	0.41	ng	97
6) bis(2-Chloroethyl)ether	7.59	93	591	0.28	ng	86
9) Naphthalene	10.97	128	2963	0.35	ng	# 95
10) Hexachlorobutadiene	11.26	225	446	0.35	ng	99
11) 2-Methylnaphthalene	12.63	142	1786	0.33	ng	96
15) Acenaphthylene	14.51	152	13648	0.32	ng	100
16) Acenaphthene	14.87	154	2293	0.33	ng	96
17) Fluorene	15.86	166	2784	0.33	ng	97
19) Hexachlorobenzene	16.87	284	1030	0.39	ng	# 82
20) Pentachlorophenol	17.23	266	159	0.40	ng	92
21) Phenanthrene	17.58	178	5203	0.35	ng	# 78
22) Anthracene	17.69	178	4726	0.34	ng	99
23) Fluoranthene	19.59	202	22989	0.33	ng	99
25) Pyrene	19.95	202	24177	0.32	ng	100
27) Benzo(a)anthracene	21.70	228	54055	0.68	ng	99
28) Chrysene	21.70	228	54377	0.67	ng	# 79
29) Bis(2-ethylhexyl)phthalate	21.61	149	1875	0.42	ng	# 75
30) Indeno(1,2,3-cd)pyrene	26.56	276	26124	0.34	ng	98
32) Benzo(b)fluoranthene	23.36	252	6294	0.32	ng	97
33) Benzo(k)fluoranthene	23.40	252	6320	0.31	ng	99
34) Benzo(a)pyrene	23.97	252	5410	0.30	ng	98
35) Dibenzo(a,h)anthracene	26.60	278	5242	0.29	ng	97
36) Benzo(g,h,i)perylene	27.33	276	23077	0.30	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE120116\
Data File : BE092587.D
Acq On : 1 Dec 2016 14:36
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Dec 02 01:01:18 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 15:43:18 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. -0.02 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion: 152 Resp: 8138

Ion Ratio Lower Upper

152 100

150 183.7 106.0 159.0#




115 91.7 31.2 46.8#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.41 ng

RT: 3.39 min Scan# 69

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion: 88 Resp: 498

Ion Ratio Lower Upper

88 100

58 66.5 63.6 95.4



43 33.1 28.0 42.0




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.41 ng

RT: 3.79 min Scan# 104

Delta R.T. 0.01 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion: 42 Resp: 332

Ion Ratio Lower Upper

42 100

74 104.2 86.0 129.0




44 6.9 5.1 7.7



Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 0.34 ng

RT: 5.68 min Scan# 365

Delta R.T. 0.01 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion: 112 Resp: 474

Ion Ratio Lower Upper

112 100

64 61.6 53.5 80.3



63 34.6 27.9 41.9



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 0.38 ng

RT: 7.36 min Scan# 541

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion: 99 Resp: 614

Ion Ratio Lower Upper

99 100

42 0.0 16.0 24.0#


71 33.1 30.6 45.8





Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



Time-->



#6

bis(2-Chloroethyl)ether

Concen: 0.28 ng

RT: 7.59 min Scan# 551

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion: 93 Resp: 591

Ion Ratio Lower Upper

93 100

63 71.7 71.6 107.4


95 30.6 24.0 36.0




Abundance Ion 93.00 (92.70 to 93.70): BE0


Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0



Time-->





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.93 min Scan# 707

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion: 136 Resp: 41115



Ion Ratio Lower Upper

136 100

137 11.5 8.2 12.4

54 7.2 9.6 14.4#

68 5.2 6.7 10.1#





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.35 ng

RT: 9.36 min Scan# 628

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36



Tgt Ion: 82 Resp: 723

Ion Ratio Lower Upper

82 100

128 53.2 39.0 58.4


54 67.6 44.8 67.2#




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0









#9
Naphthalene
Concen: 0.35 ng
RT: 10.97 min Scan# 709
Delta R.T. -0.02 min
Lab File: BE092587.D
Acq: 1 Dec 2016 14:36




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:128 Resp: 2963
Ion Ratio Lower Upper
128 100
129 14.4 11.2 16.8
127 17.9 11.6 17.4#



#10
Hexachlorobutadiene
Concen: 0.35 ng
RT: 11.26 min Scan# 724
Delta R.T. 0.00 min
Lab File: BE092587.D
Acq: 1 Dec 2016 14:36

Tgt Ion:225 Resp: 446
Ion Ratio Lower Upper
225 100
223 62.3 50.6 76.0
227 63.2 51.4 77.0





#11

2-Methylnaphthalene

Concen: 0.33 ng

RT: 12.63 min Scan# 815

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion:142 Resp: 1786

Ion Ratio Lower Upper

142 100

141 87.2 73.7 110.5




115 41.1 34.0 51.0



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.80 min Scan# 979

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion:164 Resp: 30160

Ion Ratio Lower Upper

164 100

162 100.0 71.4 107.0



160 43.6 27.4 41.2#




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 0.30 ng
 RT: 16.31 min Scan# 1088
 Delta R.T. -0.00 min
 Lab File: BE092587.D
 Acq: 1 Dec 2016 14:36




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion: 330 Resp: 203
 Ion Ratio Lower Upper
 330 100
 332 97.0 75.4 113.0
 141 41.4 31.0 46.6



#14
 2-Fluorobiphenyl
 Concen: 0.33 ng
 RT: 13.43 min Scan# 884
 Delta R.T. -0.00 min
 Lab File: BE092587.D
 Acq: 1 Dec 2016 14:36

Tgt Ion: 172 Resp: 2410
 Ion Ratio Lower Upper
 172 100
 171 36.7 30.1 45.1
 170 25.5 21.8 32.6





#15

Acenaphthylene

Concen: 0.32 ng

RT: 14.51 min Scan# 962

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:152 Resp: 13648

Ion Ratio Lower Upper

152 100

151 4.8 3.9 5.9

153 13.0 10.4 15.6



#16

Acenaphthene

Concen: 0.33 ng

RT: 14.87 min Scan# 983

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36




Tgt Ion:154 Resp: 2293


Ion Ratio Lower Upper

154 100

153 437.9 350.8 526.2

152 233.3 171.1 256.7





#17

Fluorene

Concen: 0.33 ng

RT: 15.86 min Scan# 1049

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion:166 Resp: 2784

Ion Ratio Lower Upper

166 100

165 100.6 78.2 117.4




167 13.3 11.0 16.4



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.56 min Scan# 1153

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion:188 Resp: 66559

Ion Ratio Lower Upper

188 100

94 4.0 3.2 4.8



80 3.6 2.6 3.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0









#19
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.87 min Scan# 1123
Delta R.T. -0.00 min
Lab File: BE092587.D
Acq: 1 Dec 2016 14:36




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 284 Resp: 1030
Ion Ratio Lower Upper
284 100
142 53.0 29.1 43.7#
249 39.2 27.9 41.9



#20
Pentachlorophenol
Concen: 0.40 ng
RT: 17.23 min Scan# 1139
Delta R.T. 0.00 min
Lab File: BE092587.D
Acq: 1 Dec 2016 14:36

Tgt Ion: 266 Resp: 159
Ion Ratio Lower Upper
266 100
268 81.2 62.5 93.7
264 82.4 57.2 85.8





#21

Phenanthrene

Concen: 0.35 ng

RT: 17.58 min Scan# 1154

Delta R.T. -0.02 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion:178 Resp: 5203

Ion Ratio Lower Upper

178 100

176 19.7 15.8 23.8




179 0.0 16.0 24.0#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#22

Anthracene

Concen: 0.34 ng

RT: 17.69 min Scan# 1159

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion:178 Resp: 4726

Ion Ratio Lower Upper

178 100

176 18.6 15.0 22.4



179 16.0 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#23

Fluoranthene

Concen: 0.33 ng

RT: 19.59 min Scan# 1265

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion:202 Resp: 22989

Ion Ratio Lower Upper

202 100

101 2.9 2.2 3.4




203 17.7 13.7 20.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.72 min Scan# 1374

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion:240 Resp: 95773

Ion Ratio Lower Upper

240 100

120 2.6 2.6 4.0



236 27.4 24.6 37.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#25
 Pyrene
 Concen: 0.32 ng
 RT: 19.95 min Scan# 1286
 Delta R.T. 0.00 min
 Lab File: BE092587.D
 Acq: 1 Dec 2016 14:36




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion: 202 Resp: 24177
 Ion Ratio Lower Upper
 202 100
 200 5.3 4.2 6.4
 203 17.3 13.9 20.9



#26
 Terphenyl-d14
 Concen: 0.32 ng
 RT: 20.17 min Scan# 1299
 Delta R.T. 0.00 min
 Lab File: BE092587.D
 Acq: 1 Dec 2016 14:36

Tgt Ion: 244 Resp: 3501
 Ion Ratio Lower Upper
 244 100
 212 9.8 6.7 10.1
 122 7.1 3.6 5.4#





#27

Benzo(a)anthracene

Concen: 0.68 ng

RT: 21.70 min Scan# 1373

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion:228 Resp: 54055

Ion Ratio Lower Upper

228 100

226 29.6 23.6 35.4




229 17.3 13.3 19.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#28

Chrysene

Concen: 0.67 ng

RT: 21.70 min Scan# 1373

Delta R.T. -0.07 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion:228 Resp: 54377

Ion Ratio Lower Upper

228 100

226 29.6 36.3 54.5#



229 17.3 10.6 16.0#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.42 ng

RT: 21.61 min Scan# 1369

Delta R.T. -0.02 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tgt Ion:149 Resp: 1875

Ion Ratio Lower Upper

149 100

167 22.8 16.0 24.0




279 0.0 16.0 24.0#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.34 ng

RT: 26.56 min Scan# 1714

Delta R.T. 0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36


Tgt Ion:276 Resp: 26124

Ion Ratio Lower Upper

276 100

138 27.1 20.3 30.5



277 24.7 19.7 29.5




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#31
Perylene-d12
 Concen: 5.00 ng
 RT: 24.08 min Scan# 1543
 Delta R.T. -0.01 min
 Lab File: BE092587.D
 Acq: 1 Dec 2016 14:36




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion: 264 Resp: 80993
 Ion Ratio Lower Upper
 264 100
 260 26.0 20.1 30.1
 265 20.9 17.9 26.9



#32
Benzo(b)fluoranthene
 Concen: 0.32 ng
 RT: 23.36 min Scan# 1481
 Delta R.T. -0.00 min
 Lab File: BE092587.D
 Acq: 1 Dec 2016 14:36

Tgt Ion: 252 Resp: 6294
 Ion Ratio Lower Upper
 252 100
 253 25.1 18.7 28.1
 125 11.7 10.5 15.7





#33

Benzo(k)fluoranthene

Concen: 0.31 ng

RT: 23.40 min Scan# 1485

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 6320

Ion Ratio Lower Upper

252 100

253 24.6 20.3 30.5

125 12.0 9.6 14.4



#34

Benzo(a)pyrene

Concen: 0.30 ng

RT: 23.97 min Scan# 1534

Delta R.T. -0.00 min

Lab File: BE092587.D

Acq: 1 Dec 2016 14:36




Tgt Ion:252 Resp: 5410


Ion Ratio Lower Upper

252 100

253 24.9 19.0 28.6

125 14.1 11.8 17.8









#35
Dibenzo(a,h)anthracene
Concen: 0.29 ng
RT: 26.60 min Scan# 1716
Delta R.T. -0.00 min
Lab File: BE092587.D
Acq: 1 Dec 2016 14:36

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 278 Resp: 5242
Ion Ratio Lower Upper
278 100
139 19.6 13.8 20.8
279 27.7 21.4 32.2



#36
Benzo(g,h,i)perylene
Concen: 0.30 ng
RT: 27.33 min Scan# 1759
Delta R.T. -0.00 min
Lab File: BE092587.D
Acq: 1 Dec 2016 14:36

Tgt Ion: 276 Resp: 23077
Ion Ratio Lower Upper
276 100
277 25.7 19.8 29.8
138 26.1 19.8 29.6

