

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120116\
 Data File : BE092580.D
 Acq On : 1 Dec 2016 10:11
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 02 13:00:57 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	-0.02
2	1,4-Dioxane	0.681	0.719	-5.6	131	0.00
3	n-Nitrosodimethylamine	0.654	0.463	29.2#	99	0.01
4 S	2-Fluorophenol	0.844	0.699	17.2	99	0.00
5 S	Phenol-d6	1.104	0.845	23.5	99	0.00
6	bis(2-Chloroethyl)ether	1.341	1.231	8.2	123	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.269	0.211	21.6	95	0.00
9	Naphthalene	1.032	0.880	14.7	101	0.00
10	Hexachlorobutadiene	0.157	0.133	15.3	99	0.00
11	2-Methylnaphthalene	0.658	0.533	19.0	101	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
13 S	2,4,6-Tribromophenol	0.112	0.081	27.7#	94	0.00
14 S	2-Fluorobiphenyl	1.219	1.006	17.5	100	0.00
15	Acenaphthylene	6.962	5.562	20.1	101	0.00
16	Acenaphthene	1.136	0.969	14.7	103	0.00
17	Fluorene	1.408	1.149	18.4	100	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
19	Hexachlorobenzene	0.197	0.160	18.8	104	0.00
20	Pentachlorophenol	0.036	0.026	27.8#	106	0.02
21	Phenanthrene	1.119	0.951	15.0	101	0.00
22	Anthracene	1.053	0.869	17.5	101	0.00
23	Fluoranthene	5.219	4.290	17.8	101	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
25	Pyrene	3.975	3.162	20.5	100	0.00
26 S	Terphenyl-d14	0.572	0.461	19.4	103	0.00
27	Benzo(a)anthracene	4.692	3.434	26.8#	102	0.00
28	Chrysene	4.250	3.778	11.1	101	-0.02
29	Bis(2-ethylhexyl)phthalate	0.305	0.213	30.2#	98	-0.02
30	Indeno(1,2,3-cd)pyrene	4.722	3.331	29.5#	87	0.00
31 I	Perylene-d12	1.000	1.000	0.0	114	-0.01
32	Benzo(b)fluoranthene	1.213	0.953	21.4	93	0.00
33	Benzo(k)fluoranthene	1.270	0.944	25.7#	89	0.00
34 C	Benzo(a)pyrene	1.122	0.816	27.3#	89	0.00
35	Dibenzo(a,h)anthracene	1.111	0.785	29.3#	86	0.00
36	Benzo(g,h,i)perylene	4.702	3.519	25.2#	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1