

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120517\
 Data File : BE094930.D
 Acq On : 6 Dec 2017 00:23
 Operator : SJ/JU
 Sample : I6698-11
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FT-MW02S-20171130

Quant Time: Dec 06 04:06:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.48	152	9999	0.40	ng	0.00
7) Naphthalene-d8	11.33	136	24276	0.40	ng	0.00
13) Acenaphthene-d10	15.08	164	33839	0.40	ng	0.00
19) Phenanthrene-d10	17.77	188	26739	0.40	ng	0.00
27) Chrysene-d12	21.89	240	32488	0.40	ng	0.00
34) Perylene-d12	24.57	264	21936	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.95	112	3288	0.21	ng	0.00
5) Phenol-d6	7.60	99	2509	0.11	ng	0.00
8) Nitrobenzene-d5	9.66	82	7899	0.55	ng	0.00
11) 2-Methylnaphthalene-d10	12.86	152	16538	0.43	ng	0.00
14) 2,4,6-Tribromophenol	16.55	330	1725	0.30	ng	0.02
15) 2-Fluorobiphenyl	13.71	172	22779	0.16	ng	0.00
25) Fluoranthene-d10	19.76	212	128645	0.35	ng	0.00
29) Terphenyl-d14	20.33	244	22114	0.37	ng	0.00


Target Compounds				Qvalue		
3) n-Nitrosodimethylamine	4.01	42	850	0.064	ng	# 1
9) Naphthalene	11.38	128	7593	0.027	ng	# 38
16) Acenaphthylene	14.81	152	5067	0.027	ng	# 1
17) Acenaphthene	15.18	154	6181	0.050	ng	# 61
18) Fluorene	16.16	166	5266	0.033	ng	# 10
22) Pentachlorophenol	17.52	266	31	0.140	ng	94
23) Phenanthrene	17.74	178	2697	0.032	ng	# 61
24) Anthracene	17.74	178	2697	0.030	ng	# 62
28) Pyrene	20.14	202	10664	0.095	ng	97
30) Benzo(a)anthracene	21.87	228	3178	0.032	ng	# 87
31) Chrysene	21.87	228	3178	0.029	ng	# 87
32) Bis(2-ethylhexyl)phthalate	21.76	149	60987	0.369	ng	# 92
33) Indeno(1,2,3-cd)pyrene	27.43	276	3981	0.044	ng	# 85
35) Benzo(b)fluoranthene	23.79	252	3392	0.042	ng	87
36) Benzo(k)fluoranthene	23.79	252	3392	0.042	ng	87
37) Benzo(a)pyrene	24.45	252	2415	0.034	ng	92
38) Dibenzo(a,h)anthracene	27.46	278	1974	0.030	ng	# 89
39) Benzo(a,h,i)perylene	28.33	276	3792	0.059	ng	# 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120517\
Data File : BE094930.D
Acq On : 6 Dec 2017 00:23
Operator : SJ/JU
Sample : I6698-11
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FT-MW02S-20171130

Quant Time: Dec 06 04:06:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 8.48 min Scan# 447
Delta R.T. -0.00 min
Lab File: BE094930.D
Acq: 6 Dec 2017 00:23

Instrument :
BNA_E
ClientSampleId :
FT-MW02S-20171130

Tot Ion:152 Resp: 9999
Ion Ratio Lower Upper
152 100
150 135.9 117.2 175.8
115 64.4 57.0 85.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#3


n-Nitrosodimethylamine
Concen: 0.064 ng
RT: 4.01 min Scan# 92
Delta R.T. -0.02 min
Lab File: BE094930.D
Acq: 6 Dec 2017 00:23

Tgt Ion: 42 Resp: 850
Ion Ratio Lower Upper
42 100
74 10.1 100.3 150.5#
44 76.8 17.0 25.4#



Abundance Ion 42.00 (41.70 to 42.70): BE0
Ion 74.00 (73.70 to 74.70): BE0
Ion 44.00 (43.70 to 44.70): BE0





#4

2-Fluorophenol

Concen: 0.208 ng

RT: 5.95 min Scan# 246

Delta R.T. -0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:112 Resp: 3288

Ion Ratio Lower Upper

112 100

64 78.5 60.1 90.1




63 118.6 116.8 175.2



Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#5

Phenol-d6

Concen: 0.107 ng

RT: 7.60 min Scan# 377

Delta R.T. 0.01 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion: 99 Resp: 2509

Ion Ratio Lower Upper

99 100

42 149.2 20.2 30.2#



71 0.0 28.4 42.6#




Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 11.33 min Scan# 613

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130

Tot Ion:136 Resp: 24276


Ion Ratio Lower Upper

136 100

137 15.2 9.5 14.3#

54 164.5 29.1 43.7#

68 50.6 6.2 9.2#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.550 ng

RT: 9.66 min Scan# 514

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion: 82 Resp: 7899

Ion Ratio Lower Upper

82 100

128 112.4 150.0 225.0#



54 210.3 154.2 231.4




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.027 ng

RT: 11.38 min Scan# 616

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130





Tot Ion:128 Resp: 7593

Ion Ratio Lower Upper

128 100

129 14.8 2.6 3.8#

127 32.2 2.8 4.2#



#11

2-Methylnaphthalene-d10

Concen: 0.427 ng

RT: 12.86 min Scan# 845

Delta R.T. -0.01 min

Lab File: BE094930.D




Acq: 6 Dec 2017 00:23


Tgt Ion:152 Resp: 16538

Ion Ratio Lower Upper

152 100

151 27.6 15.4 23.0#





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 15.08 min Scan# 1005

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:164 Resp: 33839

Ion Ratio Lower Upper




164 100

162 52.5 83.1 124.7#

160 38.0 37.1 55.7



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.296 ng

RT: 16.55 min Scan# 1101

Delta R.T. 0.02 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion:330 Resp: 1725

Ion Ratio Lower Upper



330 100


332 98.0 152.7 229.1#

141 0.0 33.7 50.5#



Abundance Ion 330.00 (329.70 to 330.70): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E





#15

2-Fluorobiphenyl

Concen: 0.157 ng

RT: 13.71 min Scan# 913

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:172 Resp: 22779

Ion Ratio Lower Upper

172 100

171 33.9 29.9 44.9




170 24.1 21.8 32.8



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#16

Acenaphthylene

Concen: 0.027 ng

RT: 14.81 min Scan# 987

Delta R.T. 0.02 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion:152 Resp: 5067

Ion Ratio Lower Upper

152 100

151 567.5 15.7 23.5#



153 156.6 10.5 15.7#




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E





#17

Acenaphthene

Concen: 0.050 ng

RT: 15.18 min Scan# 1012

Delta R.T. 0.05 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130





Tot Ion:154 Resp: 6181

Ion Ratio Lower Upper

154 100

153 134.7 89.2 133.8#

152 0.0 42.0 63.0#



#18

Fluorene

Concen: 0.033 ng

RT: 16.16 min Scan# 1075

Delta R.T. 0.06 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23




Tgt Ion:166 Resp: 5266


Ion Ratio Lower Upper

166 100

165 0.0 77.8 116.6#

167 0.0 42.4 63.6#





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.77 min Scan# 1183

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130





Tot Ion:188 Resp: 26739

Ion Ratio Lower Upper

188 100

94 0.0 3.9 5.9#

80 142.0 3.6 5.4#



#22

Pentachlorophenol

Concen: 0.140 ng

RT: 17.52 min Scan# 1166

Delta R.T. 0.09 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23




Tgt Ion:266 Resp: 31


Ion Ratio Lower Upper

266 100

264 91.8 72.5 108.7

268 103.1 73.8 110.6





#23

Phenanthrene

Concen: 0.032 ng

RT: 17.74 min Scan# 1181

Delta R.T. -0.07 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:178 Resp: 2697

Ion Ratio Lower Upper

178 100

176 0.0 15.1 22.7#




179 0.0 11.8 17.6#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#24

Anthracene

Concen: 0.030 ng

RT: 17.74 min Scan# 1181

Delta R.T. -0.16 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion:178 Resp: 2697

Ion Ratio Lower Upper

178 100

176 0.0 12.8 19.2#



179 0.0 13.0 19.6#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#25
 Fluoranthene-d10
 Concen: 0.352 ng
 RT: 19.76 min Scan# 1432
 Delta R.T. 0.00 min
 Lab File: BE094930.D
 Acq: 6 Dec 2017 00:23




Instrument :
 BNA_E
 ClientSampleId :
 FT-MW02S-20171130


Tot Ion:212 Resp: 128645
 Ion Ratio Lower Upper
 212 100
 106 4.0 2.8 4.2
 104 2.5 1.6 2.4#



#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.89 min Scan# 1583
 Delta R.T. 0.00 min
 Lab File: BE094930.D
 Acq: 6 Dec 2017 00:23

Tgt Ion:240 Resp: 32488
 Ion Ratio Lower Upper
 240 100
 120 18.2 5.5 8.3#
 236 29.5 20.7 31.1





#28

Pvrene

Concen: 0.095 ng

RT: 20.14 min Scan# 1471

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130





Tot Ion:202 Resp: 10664

Ion Ratio Lower Upper

202 100

200 22.4 16.6 25.0

203 15.9 13.8 20.8



#29

Terphenyl-d14

Concen: 0.369 ng

RT: 20.33 min Scan# 1483

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23




Tgt Ion:244 Resp: 22114


Ion Ratio Lower Upper

244 100

212 14.8 25.4 38.0#

122 42.4 8.1 12.1#





#30

Benzo(a)anthracene

Concen: 0.032 ng

RT: 21.87 min Scan# 1582

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:228 Resp: 3178

Ion Ratio Lower Upper

228 100

226 53.2 40.0 60.0




229 64.2 40.0 60.0#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#31

Chrysene

Concen: 0.029 ng

RT: 21.87 min Scan# 1582

Delta R.T. -0.05 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion:228 Resp: 3178

Ion Ratio Lower Upper

228 100

226 53.2 40.0 60.0



229 64.2 40.0 60.0#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#32

Bis(2-ethylhexyl)phthalate

Concen: 0.369 ng

RT: 21.76 min Scan# 1575

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:149 Resp: 60987

Ion Ratio Lower Upper

149 100

167 9.6 5.4 8.0#




279 0.7 0.7 1.1#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#33

Indeno(1,2,3-cd)pyrene

Concen: 0.044 ng

RT: 27.43 min Scan# 2774

Delta R.T. -0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion:276 Resp: 3981

Ion Ratio Lower Upper

276 100

138 13.4 22.5 33.7#



277 24.7 19.9 29.9




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#34
 Pervlene-d12
 Concen: 0.400 ng
 RT: 24.57 min Scan# 2148
 Delta R.T. 0.00 min
 Lab File: BE094930.D
 Acq: 6 Dec 2017 00:23




Instrument :
 BNA_E
 ClientSampleId :
 FT-MW02S-20171130


Tot Ion:264 Resp: 21936
 Ion Ratio Lower Upper
 264 100
 260 25.8 20.5 30.7
 265 34.1 22.8 34.2



#35
 Benzo(b)fluoranthene
 Concen: 0.042 ng
 RT: 23.79 min Scan# 1977
 Delta R.T. 0.06 min
 Lab File: BE094930.D
 Acq: 6 Dec 2017 00:23

Tgt Ion:252 Resp: 3392
 Ion Ratio Lower Upper
 252 100
 253 50.6 48.0 72.0
 125 58.7 56.0 84.0





#36

Benzo(k)fluoranthene

Concen: 0.042 ng

RT: 23.79 min Scan# 1977

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:252 Resp: 3392

Ion Ratio Lower Upper

252 100

253 50.6 48.0 72.0




125 58.7 56.0 84.0



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#37

Benzo(a)pyrene

Concen: 0.034 ng

RT: 24.45 min Scan# 2121

Delta R.T. 0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion:252 Resp: 2415

Ion Ratio Lower Upper

252 100

253 58.5 52.0 78.0



125 77.7 68.0 102.0




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#38

Dibenzo(a,h)anthracene

Concen: 0.030 ng

RT: 27.46 min Scan# 2780

Delta R.T. -0.01 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23

Instrument :

BNA_E

ClientSampleId :

FT-MW02S-20171130


Tot Ion:278 Resp: 1974

Ion Ratio Lower Upper

278 100

139 72.8 56.0 84.0




279 49.6 52.0 78.0#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#39

Benzo(g,h,i)perylene

Concen: 0.059 ng

RT: 28.33 min Scan# 2970

Delta R.T. -0.00 min

Lab File: BE094930.D

Acq: 6 Dec 2017 00:23


Tgt Ion:276 Resp: 3792

Ion Ratio Lower Upper

276 100

277 33.7 52.0 78.0#

138 39.5 44.0 66.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

