

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120517\
 Data File : BE094929.D
 Acq On : 5 Dec 2017 23:47
 Operator : SJ/JU
 Sample : I6698-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FT-DUP02-20171130

Manual Integrations
 APPROVED

Sohil
 12/6/2017 5:10:28 PM

Quant Time: Dec 06 06:51:51 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.49	152	8274	0.40	ng	0.00
7) Naphthalene-d8	11.33	136	17947	0.40	ng	0.00
13) Acenaphthene-d10	15.06	164	10260	0.40	ng	-0.02
19) Phenanthrene-d10	17.77	188	26029	0.40	ng	0.00
27) Chrysene-d12	21.88	240	27532	0.40	ng	0.00
34) Perylene-d12	24.56	264	22417	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.96	112	1412	0.11	ng	0.00
5) Phenol-d6	7.59	99	1402	0.07	ng	0.00
8) Nitrobenzene-d5	9.66	82	4464	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	12.86	152	11542	0.40	ng	0.00
14) 2,4,6-Tribromophenol	16.53	330	446	0.27	ng	0.00
15) 2-Fluorobiphenyl	13.71	172	19778	0.45	ng	0.00
25) Fluoranthene-d10	19.76	212	143687	0.40	ng	0.00
29) Terphenyl-d14	20.33	244	30517	0.60	ng	0.00
Target Compounds						
30) Benzo(a)anthracene	21.85	228	2526	0.030	ng	Qvalue # 67
31) Chrysene	21.92	228	2620m	0.028	ng	
32) Bis(2-ethylhexyl)phthalate	21.74	149	12176	0.087	ng	100
33) Indeno(1,2,3-cd)pyrene	27.42	276	2250	0.029	ng	# 73
35) Benzo(b)fluoranthene	23.73	252	2582	0.032	ng	# 60
36) Benzo(k)fluoranthene	23.78	252	2239	0.027	ng	# 61
37) Benzo(a)pyrene	24.44	252	2050	0.029	ng	# 59
38) Dibenzo(a,h)anthracene	27.46	278	1862	0.028	ng	# 60
39) Benzo(g,h,i)perylene	28.32	276	2304	0.035	ng	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120517\
Data File : BE094929.D
Acq On : 5 Dec 2017 23:47
Operator : SJ/JU
Sample : I6698-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FT-DUP02-20171130

Manual Integrations
APPROVED

Sohil
12/6/2017 5:10:28 PM

Quant Time: Dec 06 06:51:51 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

