

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE120517\
 Data File : BE094915.D
 Acq On : 5 Dec 2017 15:27
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 06 04:02:24 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	85	0.00
2	1,4-Dioxane	0.400	0.391	2.3	78	0.00
3	n-Nitrosodimethylamine	0.400	0.387	3.3	83	0.00
4 S	2-Fluorophenol	0.400	0.219	45.3#	46	0.00
5 S	Phenol-d6	0.400	0.274	31.5#	58	0.00
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	82	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.230	42.5#	49	0.00
9	Naphthalene	0.400	0.403	-0.8	85	0.00
10	Hexachlorobutadiene	0.400	0.410	-2.5	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.394	1.5	83	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	84	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.254	36.5#	40	0.00
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	84	0.00
16	Acenaphthylene	0.400	0.376	6.0	81	0.00
17	Acenaphthene	0.400	0.388	3.0	83	0.00
18	Fluorene	0.400	0.384	4.0	80	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.02
20	4-Bromophenyl-phenylether	0.400	0.381	4.8	82	0.02
21	Hexachlorobenzene	0.400	0.415	-3.7	87	0.02
22	Pentachlorophenol	0.400	0.283	29.3#	47	0.00
23	Phenanthrene	0.400	0.397	0.8	83	0.00
24	Anthracene	0.400	0.366	8.5	80	0.00
25 SURR	Fluoranthene-d10	0.400	0.357	10.8	76	0.00
26	Fluoranthene	0.400	0.369	7.8	78	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	73	0.00
28	Pvrene	0.400	0.411	-2.7	76	0.00
29 S	Terphenyl-d14	0.400	0.402	-0.5	75	0.00
30	Benzo(a)anthracene	0.400	0.370	7.5	70	0.00
31	Chrysene	0.400	0.388	3.0	71	0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.113	71.8#	24	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.324	19.0	61	0.00
34 I	Perylene-d12	0.400	0.400	0.0	65	0.00
35	Benzo(b)fluoranthene	0.400	0.395	1.3	66	0.00
36	Benzo(k)fluoranthene	0.400	0.390	2.5	66	0.00
37 C	Benzo(a)pyrene	0.400	0.379	5.3	64	0.00
38	Dibenzo(a,h)anthracene	0.400	0.355	11.3	60	0.00
39	Benzo(a,h,i)perylene	0.400	0.385	3.8	64	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE120517\
Data File : BE094915.D
Acq On : 5 Dec 2017 15:27
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 06 04:02:24 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			