

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE120517\
 Data File : BE094933.D
 Acq On : 6 Dec 2017 2:11
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 06 06:56:27 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	-0.01
2	1,4-Dioxane	0.419	0.435	-3.8	77	-0.01
3	n-Nitrosodimethylamine	0.533	0.521	2.3	78	-0.01
4 S	2-Fluorophenol	0.633	0.692	-9.3	85	0.00
5 S	Phenol-d6	0.938	1.038	-10.7	87	0.00
6	bis(2-Chloroethyl)ether	0.930	0.881	5.3	73	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
8 S	Nitrobenzene-d5	0.237	0.288	-21.5	97	0.00
9	Naphthalene	4.690	4.603	1.9	77	0.00
10	Hexachlorobutadiene	0.198	0.208	-5.1	82	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.655	-2.7	81	0.00
12	2-Methylnaphthalene	1.165	1.182	-1.5	81	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	-0.01
14 S	2,4,6-Tribromophenol	0.086	0.126	-46.5#	129	0.00
15 S	2-Fluorobiphenyl	1.713	1.670	2.5	82	0.00
16	Acenaphthylene	2.260	2.188	3.2	85	0.00
17	Acenaphthene	1.461	1.424	2.5	85	-0.01
18	Fluorene	1.859	1.786	3.9	81	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.223	0.221	0.9	85	0.00
21	Hexachlorobenzene	0.218	0.225	-3.2	86	0.00
22	Pentachlorophenol	0.054	0.069	-27.8#	120	-0.01
23	Phenanthrene	1.275	1.225	3.9	81	0.00
24	Anthracene	1.351	1.268	6.1	82	0.00
25 SURR	Fluoranthene-d10	5.466	5.162	5.6	81	0.00
26	Fluoranthene	1.638	1.507	8.0	78	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
28	Pvrene	1.381	1.645	-19.1	87	-0.02
29 S	Terphenyl-d14	0.739	0.795	-7.6	79	-0.02
30	Benzo(a)anthracene	1.227	1.302	-6.1	79	-0.02
31	Chrysene	1.355	1.265	6.6	68	0.00
32	Bis(2-ethylhexyl)phthalate	2.033	2.543	-25.1#	106	-0.02
33	Indeno(1,2,3-cd)pyrene	1.109	1.159	-4.5	77	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	78	0.00
35	Benzo(b)fluoranthene	1.460	1.380	5.5	76	0.00
36	Benzo(k)fluoranthene	1.489	1.387	6.9	75	0.00
37 C	Benzo(a)pyrene	1.280	1.259	1.6	79	0.00
38	Dibenzo(a,h)anthracene	1.183	1.129	4.6	76	-0.02
39	Benzo(a,h,i)perylene	1.181	1.139	3.6	77	-0.02

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE120517\
Data File : BE094933.D
Acq On : 6 Dec 2017 2:11
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 06 06:56:27 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			