

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120619\
 Data File : BE100799.D
 Acq On : 6 Dec 2019 9:55
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 06 12:49:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.729	0.637	12.6	97	-0.01
3	n-Nitrosodimethylamine	0.594	0.586	1.3	126	-0.01
4 S	2-Fluorophenol	0.949	1.148	-21.0	143	0.00
5 S	Phenol-d6	1.433	1.768	-23.4	148	0.00
6	bis(2-Chloroethyl)ether	1.385	1.376	0.6	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
8 S	Nitrobenzene-d5	0.190	0.282	-48.4#	209#	0.00
9	Naphthalene	4.250	4.072	4.2	123	0.00
10	Hexachlorobutadiene	0.166	0.158	4.8	121	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.571	1.6	129	0.00
12	2-Methylnaphthalene	0.664	0.645	2.9	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
14 S	2,4,6-Tribromophenol	0.044	0.089	-102.3#	305#	0.00
15 S	2-Fluorobiphenyl	1.587	1.432	9.8	126	0.00
16	Acenaphthylene	1.655	1.593	3.7	149	0.00
17	Acenaphthene	1.140	1.061	6.9	136	0.00
18	Fluorene	1.471	1.321	10.2	135	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	-0.01
20	4-Bromophenyl-phenylether	0.196	0.194	1.0	130	0.00
21	Hexachlorobenzene	0.213	0.211	0.9	125	0.00
22	Pentachlorophenol	0.015	0.032	-113.3#	1314#	0.00
23	Phenanthrene	1.212	1.135	6.4	118	0.00
24	Anthracene	1.013	0.971	4.1	133	0.00
25 SURR	Fluoranthene-d10	4.728	4.131	12.6	115	0.00
26	Fluoranthene	1.446	1.229	15.0	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.300	1.323	-1.8	115	0.00
29 S	Terphenyl-d14	0.958	0.936	2.3	108	0.00
30	Benzo(a)anthracene	1.179	1.300	-10.3	131	0.00
31	Chrysene	1.385	1.321	4.6	104	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	2.523	-136.0#	329#	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.485	-23.0	147	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	126	0.00
35	Benzo(b)fluoranthene	1.095	1.066	2.6	129	0.00
36	Benzo(k)fluoranthene	1.228	1.094	10.9	118	0.00
37 C	Benzo(a)pyrene	0.948	0.985	-3.9	143	0.00
38	Dibenzo(a,h)anthracene	0.943	0.973	-3.2	146	-0.02
39	Benzo(a,h,i)perylene	1.009	1.006	0.3	133	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120619\
Data File : BE100799.D
Acq On : 6 Dec 2019 9:55
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 06 12:49:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			