

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120619\
 Data File : BE100803.D
 Acq On : 6 Dec 2019 12:59
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 06 13:34:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.729	0.648	11.1	96	-0.01
3	n-Nitrosodimethylamine	0.594	0.594	0.0	124	-0.01
4 S	2-Fluorophenol	0.949	1.136	-19.7	137	0.00
5 S	Phenol-d6	1.433	1.758	-22.7	143	0.00
6	bis(2-Chloroethyl)ether	1.385	1.365	1.4	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
8 S	Nitrobenzene-d5	0.190	0.276	-45.3#	197#	0.00
9	Naphthalene	4.250	4.043	4.9	118	0.00
10	Hexachlorobutadiene	0.166	0.155	6.6	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.571	1.6	124	0.00
12	2-Methylnaphthalene	0.664	0.645	2.9	122	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.044	0.090	-104.5#	298#	0.00
15 S	2-Fluorobiphenyl	1.587	1.416	10.8	121	0.00
16	Acenaphthylene	1.655	1.584	4.3	143	0.00
17	Acenaphthene	1.140	1.081	5.2	134	0.00
18	Fluorene	1.471	1.361	7.5	134	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	-0.01
20	4-Bromophenyl-phenylether	0.196	0.182	7.1	125	0.00
21	Hexachlorobenzene	0.213	0.200	6.1	121	0.00
22	Pentachlorophenol	0.015	0.029	-93.3#	1247#	0.00
23	Phenanthrene	1.212	1.099	9.3	117	0.01
24	Anthracene	1.013	0.934	7.8	131	0.00
25 SURR	Fluoranthene-d10	4.728	3.927	16.9	112	0.00
26	Fluoranthene	1.446	1.180	18.4	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	1.300	1.308	-0.6	112	0.00
29 S	Terphenyl-d14	0.958	0.923	3.7	105	0.00
30	Benzo(a)anthracene	1.179	1.252	-6.2	125	0.00
31	Chrysene	1.385	1.301	6.1	101	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	2.167	-102.7#	279#	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.410	-16.8	138	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.095	1.056	3.6	122	0.00
36	Benzo(k)fluoranthene	1.228	1.086	11.6	112	0.00
37 C	Benzo(a)pyrene	0.948	0.966	-1.9	135	0.00
38	Dibenzo(a,h)anthracene	0.943	0.962	-2.0	138	-0.02
39	Benzo(a,h,i)perylene	1.009	0.992	1.7	126	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE120619\
Data File : BE100803.D
Acq On : 6 Dec 2019 12:59
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 06 13:34:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			