

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120817\
 Data File : BE094954.D
 Acq On : 8 Dec 2017 18:23
 Operator : SJ/JU
 Sample : I6740-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FT-PZ-462S-20171202

Manual Integrations
 APPROVED

Sohil
 12/11/2017 6:18:04 PM

Quant Time: Dec 09 01:25:27 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.48	152	7597	0.40	ng	0.00
7) Naphthalene-d8	11.33	136	16704	0.40	ng	0.00
13) Acenaphthene-d10	15.08	164	9954	0.40	ng	0.00
19) Phenanthrene-d10	17.77	188	23378	0.40	ng	0.00
27) Chrysene-d12	21.88	240	23700m	0.40	ng	0.00
34) Perylene-d12	24.56	264	20200	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.95	112	537	0.04	ng	0.00
5) Phenol-d6	7.60	99	1617	0.09	ng	0.01
8) Nitrobenzene-d5	9.67	82	2442	0.25	ng	0.00
11) 2-Methylnaphthalene-d10	12.87	152	4126	0.15	ng	0.00
14) 2,4,6-Tribromophenol	16.53	330	369	0.25	ng	0.00
15) 2-Fluorobiphenyl	13.71	172	10881	0.26	ng	0.00
25) Fluoranthene-d10	19.76	212	51618	0.16	ng	0.00
29) Terphenyl-d14	20.33	244	22843	0.52	ng	0.00
Target Compounds						
22) Pentachlorophenol	17.43	266	2048	0.620	ng	Qvalue # 72
23) Phenanthrene	17.82	178	2985	0.040	ng	# 87
32) Bis(2-ethylhexyl)phthalate	21.74	149	41373	0.344	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120817\
Data File : BE094954.D
Acq On : 8 Dec 2017 18:23
Operator : SJ/JU
Sample : I6740-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FT-PZ-462S-20171202

Manual Integrations
APPROVED

Sohil
12/11/2017 6:18:04 PM

Quant Time: Dec 09 01:25:27 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

