

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120817\
 Data File : BE094955.D
 Acq On : 8 Dec 2017 18:59
 Operator : SJ/JU
 Sample : I6740-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FT-PZ-462I-20171202

Manual Integrations
 APPROVED

Sohil
 12/11/2017 6:18:06 PM

Quant Time: Dec 09 01:28:14 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.48	152	7173	0.40	ng	0.00
7) Naphthalene-d8	11.33	136	15828	0.40	ng	0.00
13) Acenaphthene-d10	15.07	164	9245	0.40	ng	0.00
19) Phenanthrene-d10	17.77	188	22491	0.40	ng	0.00
27) Chrysene-d12	21.89	240	22062m	0.40	ng	0.00
34) Perylene-d12	24.56	264	19146	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.95	112	1907	0.17	ng	0.00
5) Phenol-d6	7.60	99	2113	0.13	ng	0.01
8) Nitrobenzene-d5	9.67	82	5527	0.59	ng	0.00
11) 2-Methylnaphthalene-d10	12.87	152	9709	0.38	ng	0.00
14) 2,4,6-Tribromophenol	16.53	330	1071	0.53	ng	0.00
15) 2-Fluorobiphenyl	13.71	172	21463	0.54	ng	0.00
25) Fluoranthene-d10	19.76	212	113309	0.37	ng	0.00
29) Terphenyl-d14	20.33	244	26036	0.64	ng	0.00
Target Compounds						Qvalue
17) Acenaphthene	15.13	154	970	0.029	ng	95
18) Fluorene	16.10	166	1201	0.028	ng	# 78
23) Phenanthrene	17.82	178	4052	0.057	ng	96
26) Fluoranthene	19.79	202	2383	0.026	ng	95
32) Bis(2-ethylhexyl)phthalate	21.75	149	9793	0.087	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120817\
Data File : BE094955.D
Acq On : 8 Dec 2017 18:59
Operator : SJ/JU
Sample : I6740-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FT-PZ-462I-20171202

Manual Integrations
APPROVED

Sohil
12/11/2017 6:18:06 PM

Quant Time: Dec 09 01:28:14 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

