

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120817\
 Data File : BE094966.D
 Acq On : 9 Dec 2017 1:36
 Operator : SJ/JU
 Sample : I6740-12
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FT-PZ-460I-20171204

Manual Integrations
 APPROVED

Sohil
 12/11/2017 6:18:22 PM

Quant Time: Dec 09 03:13:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.48	152	8025	0.40	ng	0.00
7) Naphthalene-d8	11.33	136	17170	0.40	ng	0.00
13) Acenaphthene-d10	15.08	164	9997	0.40	ng	0.00
19) Phenanthrene-d10	17.77	188	23943	0.40	ng	0.00
27) Chrysene-d12	21.89	240	24553m	0.40	ng	0.00
34) Perylene-d12	24.56	264	20757	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.95	112	2287	0.18	ng	0.00
5) Phenol-d6	7.60	99	2353	0.13	ng	0.00
8) Nitrobenzene-d5	9.66	82	6479	0.64	ng	0.00
11) 2-Methylnaphthalene-d10	12.86	152	10970	0.40	ng	0.00
14) 2,4,6-Tribromophenol	16.53	330	1271	0.57	ng	0.00
15) 2-Fluorobiphenyl	13.71	172	28646	0.67	ng	0.00
25) Fluoranthene-d10	19.76	212	129144	0.39	ng	0.00
29) Terphenyl-d14	20.33	244	35860	0.79	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.67	88	31091	3.703	ng	Qvalue # 86
18) Fluorene	16.10	166	1216	0.026	ng	# 81
22) Pentachlorophenol	17.43	266	489	0.247	ng	# 79
23) Phenanthrene	17.82	178	5988	0.078	ng	93
24) Anthracene	17.91	178	2309m	0.029	ng	
26) Fluoranthene	19.79	202	4126	0.042	ng	97
28) Pyrene	20.14	202	3309	0.039	ng	# 89
30) Benzo(a)anthracene	21.87	228	4647	0.062	ng	# 65
31) Chrysene	21.92	228	5213m	0.063	ng	
32) Bis(2-ethylhexyl)phthalate	21.75	149	10712	0.086	ng	99
33) Indeno(1,2,3-cd)pyrene	27.44	276	3978	0.058	ng	99
35) Benzo(b)fluoranthene	23.73	252	4533	0.060	ng	# 50
36) Benzo(k)fluoranthene	23.78	252	3999	0.052	ng	# 50
38) Dibenzo(a,h)anthracene	27.46	278	3270	0.053	ng	# 52
39) Benzo(g,h,i)perylene	28.33	276	3506	0.057	ng	# 58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE120817\
Data File : BE094966.D
Acq On : 9 Dec 2017 1:36
Operator : SJ/JU
Sample : I6740-12
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FT-PZ-460I-20171204

Manual Integrations
APPROVED

Sohil
12/11/2017 6:18:22 PM

Quant Time: Dec 09 03:13:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

