

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE120817\
 Data File : BE094962.D
 Acq On : 8 Dec 2017 23:07
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 09 02:42:31 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.419	0.369	11.9	67	-0.02
3	n-Nitrosodimethylamine	0.533	0.468	12.2	72	-0.02
4 S	2-Fluorophenol	0.633	0.611	3.5	77	0.00
5 S	Phenol-d6	0.938	0.921	1.8	79	0.00
6	bis(2-Chloroethyl)ether	0.930	0.845	9.1	72	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
8 S	Nitrobenzene-d5	0.237	0.341	-43.9#	112	0.00
9	Naphthalene	4.690	4.651	0.8	76	0.00
10	Hexachlorobutadiene	0.198	0.207	-4.5	80	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.645	-1.1	78	0.00
12	2-Methylnaphthalene	1.165	1.155	0.9	77	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.086	0.107	-24.4	102	0.00
15 S	2-Fluorobiphenyl	1.713	1.764	-3.0	81	0.00
16	Acenaphthylene	2.260	2.117	6.3	76	0.00
17	Acenaphthene	1.461	1.394	4.6	77	0.00
18	Fluorene	1.859	1.766	5.0	74	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.00
20	4-Bromophenyl-phenylether	0.223	0.232	-4.0	79	0.00
21	Hexachlorobenzene	0.218	0.234	-7.3	79	0.01
22	Pentachlorophenol	0.054	0.063	-16.7	97	0.00
23	Phenanthrene	1.275	1.212	4.9	71	0.00
24	Anthracene	1.351	1.095	18.9	63	0.00
25 SURR	Fluoranthene-d10	5.466	5.114	6.4	71	0.00
26	Fluoranthene	1.638	1.529	6.7	71	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
28	Pvrene	1.381	1.605	-16.2	79	0.00
29 S	Terphenyl-d14	0.739	0.803	-8.7	74	0.00
30	Benzo(a)anthracene	1.227	1.312	-6.9	74	0.00
31	Chrysene	1.355	1.242	8.3	62	0.00
32	Bis(2-ethylhexyl)phthalate	2.033	2.095	-3.0	81	-0.02
33	Indeno(1,2,3-cd)pyrene	1.109	1.174	-5.9	73	0.00
34 I	Perylene-d12	1.000	1.000	0.0	73	0.00
35	Benzo(b)fluoranthene	1.460	1.403	3.9	72	0.00
36	Benzo(k)fluoranthene	1.489	1.405	5.6	71	0.00
37 C	Benzo(a)pyrene	1.280	1.246	2.7	74	0.00
38	Dibenzo(a,h)anthracene	1.183	1.158	2.1	74	0.00
39	Benzo(a,h,i)perylene	1.181	1.151	2.5	73	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE120817\
Data File : BE094962.D
Acq On : 8 Dec 2017 23:07
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 09 02:42:31 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			