

Data Path : Z:\HPCHEM1\BNA_E\Data\BE120914\
 Data File : BE089189.D
 Acq On : 9 Dec 2014 14:39
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Dec 09 15:35:35 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-ACID-BE120814.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 09 08:57:21 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2 S	2-Fluorophenol	1.282	1.395	-8.8	109	0.00
3 S	Phenol-d6	1.696	1.813	-6.9	110	0.00
4	2-Chlorophenol	1.399	1.476	-5.5	109	0.00
5 C	Phenol	2.016	2.178	-8.0	111	0.00
6	2-Methylphenol	1.613	1.666	-3.3	104	0.00
7	3+4-Methylphenols	1.370	1.509	-10.1	114	0.01
8 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
9 C	2-Nitrophenol	0.177	0.183	-3.4	110	0.00
10	2,4-Dimethylphenol	0.311	0.316	-1.6	110	0.00
11 C	2,4-Dichlorophenol	0.272	0.271	0.4	109	0.00
12 C	4-Chloro-3-methylphenol	0.283	0.287	-1.4	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.164	0.168	-2.4	108	0.00
15 C	2,4,6-Trichlorophenol	0.338	0.340	-0.6	111	0.00
16	2,4,5-Trichlorophenol	0.380	0.393	-3.4	110	0.00
17 P	2,4-Dinitrophenol	0.090	0.081	10.0	118	0.01
18 P	4-Nitrophenol	0.769	0.829	-7.8	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.084	0.088	-4.8	114	0.00
21 C	Pentachlorophenol	0.073	0.074	-1.4	114	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
23 I	Perylene-d12	1.000	1.000	0.0	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0