

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE120914\
 Data File : BE089197.D
 Acq On : 9 Dec 2014 21:31
 Operator : TP/JJ
 Sample : SST0.437
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SST0.437

Quant Time: Dec 10 01:04:26 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE120514.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 23:47:26 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	-0.02
2 I	Naphthalene-d8	1.000	1.000	0.0	121	-0.02
3	Naphthalene	0.958	0.994	-3.8	120	-0.02
4 SURR	2-Methylnaphthalene-d10	0.463	0.492	-6.3	123	-0.02
5	2-Methylnaphthalene	0.591	0.648	-9.6	126	-0.02
6 I	Acenaphthene-d10	1.000	1.000	0.0	122	-0.03
7	Acenaphthylene	7.001	7.249	-3.5	122	-0.02
8 C	Acenaphthene	4.263	4.489	-5.3	123	-0.02
9	Fluorene	1.162	1.219	-4.9	122	-0.02
10 I	Phenanthrene-d10	1.000	1.000	0.0	118	-0.02
11	Pentachlorophenol	0.071	0.051	28.2#	83	-0.01
12	Phenanthrene	4.311	4.570	-6.0	119	-0.02
13	Anthracene	4.248	4.436	-4.4	117	-0.02
14 SURR	Fluoranthene-d10	0.711	0.715	-0.6	113	-0.03
15 C	Fluoranthene	3.890	3.907	-0.4	112	-0.03
16 I	Chrysene-d12	1.000	1.000	0.0	106	-0.03
17	Pyrene	5.537	6.027	-8.8	111	-0.03
18	Benzo(a)anthracene	1.033	1.053	-1.9	103	-0.03
19	Chrysene	1.065	1.114	-4.6	108	-0.03
20 I	Perylene-d12	1.000	1.000	0.0	103	-0.03
21	Benzo(b)fluoranthene	0.989	1.042	-5.4	103	-0.03
22	Benzo(k)fluoranthene	1.160	1.251	-7.8	104	-0.03
23 C	Benzo(a)pyrene	1.040	1.084	-4.2	100	-0.03
24	Indeno(1,2,3-cd)pyrene	3.338	3.353	-0.4	99	-0.04
25	Dibenzo(a,h)anthracene	0.853	0.899	-5.4	102	-0.04
26	Benzo(g,h,i)perylene	4.128	4.182	-1.3	101	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0