

Data Path : Z:\HPCHEM1\BNA_E\Data\BE120914\
 Data File : Be089190.d
 Acq On : 9 Dec 2014 17:03
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Dec 10 00:56:23 2014
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-PAH-BE120514.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 08 00:50:53 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	-0.02
2	1,4-Dioxane	0.625	0.635	-1.6	111	-0.02
3 I	Naphthalene-d8	1.000	1.000	0.0	119	-0.02
4 S	Nitrobenzene-d5	0.367	0.397	-8.2	123	-0.02
5	Naphthalene	1.055	1.100	-4.3	120	-0.02
6	2-Methylnaphthalene	0.645	0.709	-9.9	127	-0.02
7 I	Acenaphthene-d10	1.000	1.000	0.0	124	-0.01
8 S	2-Fluorobiphenyl	1.384	1.475	-6.6	127	-0.01
9	Acenaphthylene	7.928	8.566	-8.0	128	-0.02
10	Acenaphthene	1.211	1.286	-6.2	129	-0.01
11	Fluorene	1.295	1.387	-7.1	127	-0.02
12 I	Phenanthrene-d10	1.000	1.000	0.0	119	-0.02
13	Pentachlorophenol	0.069	0.070	-1.4	131	-0.02
14	Phenanthrene	4.751	5.008	-5.4	120	-0.02
15	Anthracene	4.695	5.032	-7.2	121	-0.03
16	Fluoranthene	1.023	1.056	-3.2	115	-0.02
17 I	Chrysene-d12	1.000	1.000	0.0	109	-0.02
18	Pyrene	1.536	1.677	-9.2	114	-0.03
19 S	Terphenyl-d14	0.891	0.974	-9.3	116	-0.02
20	Benzo(a)anthracene	4.639	4.783	-3.1	111	-0.03
21	Chrysene	4.659	4.899	-5.2	109	-0.03
22	Indeno(1,2,3-cd)pyrene	4.105	4.423	-7.7	118	-0.04
23 I	Perylene-d12	1.000	1.000	0.0	109	-0.03
24	Benzo(b)fluoranthene	1.054	1.132	-7.4	112	-0.03
25	Benzo(k)fluoranthene	1.245	1.312	-5.4	110	-0.03
26 C	Benzo(a)pyrene	1.028	1.091	-6.1	114	-0.03
27	Dibenzo(a,h)anthracene	0.889	0.983	-10.6	117	-0.04
28	Benzo(g,h,i)perylene	4.280	4.559	-6.5	111	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0