

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120916\
 Data File : BE092590.D
 Acq On : 9 Dec 2016 12:55
 Operator : UM/SJ
 Sample : PB95153BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB95153BL

Quant Time: Dec 09 23:22:04 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	10102	5.00	ng	-0.02
7) Naphthalene-d8	10.93	136	41051	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	24805	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	50401	5.00	ng	0.00
24) Chrysene-d12	21.72	240	55528	5.00	ng	0.00
31) Perylene-d12	24.08	264	55199	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.65	112	18302	10.73	ng	-0.03
5) Phenol-d6	7.31	99	23316	8.19	ng	-0.05
8) Nitrobenzene-d5	9.32	82	10949	4.12	ng	-0.05
13) 2,4,6-Tribromophenol	16.30	330	3939	7.08	ng	-0.01
14) 2-Fluorobiphenyl	13.41	172	26101	4.32	ng	-0.02
26) Terphenyl-d14	20.16	244	31153	4.91	ng	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120916\
Data File : BE092590.D
Acq On : 9 Dec 2016 12:55
Operator : UM/SJ
Sample : PB95153BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB95153BL

Quant Time: Dec 09 23:22:04 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 15:43:18 2016
Response via : Initial Calibration

