

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120916\
 Data File : BE092596.D
 Acq On : 9 Dec 2016 16:42
 Operator : UM/SJ
 Sample : H5923-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-3

Quant Time: Dec 09 23:30:53 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	8264	5.00	ng	-0.02
7) Naphthalene-d8	10.93	136	36515	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	25454	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	55394	5.00	ng	0.00
24) Chrysene-d12	21.72	240	81386	5.00	ng	0.00
31) Perylene-d12	24.08	264	74348	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.65	112	7429	5.32	ng	-0.02
5) Phenol-d6	7.34	99	5542	2.46	ng	-0.02
8) Nitrobenzene-d5	9.32	82	11967	5.05	ng	-0.04
13) 2,4,6-Tribromophenol	16.28	330	6957	12.18	ng	-0.02
14) 2-Fluorobiphenyl	13.42	172	31760	5.12	ng	-0.01
26) Terphenyl-d14	20.16	244	54482	5.85	ng	-0.02
Target Compounds						
29) Bis(2-ethylhexyl)phthalate	21.61	149	640	0.24	ng	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE120916\
Data File : BE092596.D
Acq On : 9 Dec 2016 16:42
Operator : UM/SJ
Sample : H5923-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-3

Quant Time: Dec 09 23:30:53 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 15:43:18 2016
Response via : Initial Calibration

