

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
 Data File : BE100821.D
 Acq On : 10 Dec 2019 22:02
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 05:43:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	-0.01
2	1,4-Dioxane	0.400	0.376	6.0	105	-0.01
3	n-Nitrosodimethylamine	0.400	0.378	5.5	121	-0.01
4 S	2-Fluorophenol	0.400	0.454	-13.5	135	0.00
5 S	Phenol-d6	0.400	0.463	-15.8	140	0.00
6	bis(2-Chloroethyl)ether	0.400	0.388	3.0	114	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	124	0.00
8 S	Nitrobenzene-d5	0.400	0.534	-33.5#	190	0.00
9	Naphthalene	0.400	0.385	3.8	124	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	128	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	129	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	135	0.00
14 S	2,4,6-Tribromophenol	0.400	0.584	-46.0#	226	0.00
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	126	-0.01
16	Acenaphthylene	0.400	0.364	9.0	142	0.00
17	Acenaphthene	0.400	0.374	6.5	138	0.00
18	Fluorene	0.400	0.358	10.5	136	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	122	-0.01
20	4-Bromophenyl-phenylether	0.400	0.375	6.3	126	-0.01
21	Hexachlorobenzene	0.400	0.384	4.0	125	0.00
22	Pentachlorophenol	0.400	0.527	-31.8#	831	0.00
23	Phenanthrene	0.400	0.370	7.5	120	0.00
24	Anthracene	0.400	0.362	9.5	129	0.00
25 SURR	Fluoranthene-d10	0.400	0.320	20.0	108	0.00
26	Fluoranthene	0.400	0.320	20.0	108	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	99	0.00
28	Pvrene	0.400	0.413	-3.2	110	0.00
29 S	Terphenyl-d14	0.400	0.394	1.5	102	0.00
30	Benzo(a)anthracene	0.400	0.415	-3.7	116	-0.01
31	Chrysene	0.400	0.383	4.3	98	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.652	-63.0#	168	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.452	-13.0	127	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	123	-0.01
35	Benzo(b)fluoranthene	0.400	0.354	11.5	114	-0.01
36	Benzo(k)fluoranthene	0.400	0.342	14.5	110	-0.01
37 C	Benzo(a)pyrene	0.400	0.353	11.8	119	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.358	10.5	123	-0.03
39	Benzo(a,h,i)perylene	0.400	0.368	8.0	120	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
Data File : BE100821.D
Acq On : 10 Dec 2019 22:02
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 05:43:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				