

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
 Data File : BE100821.D
 Acq On : 10 Dec 2019 22:02
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 05:43:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	-0.01
2	1,4-Dioxane	0.729	0.686	5.9	105	-0.01
3	n-Nitrosodimethylamine	0.594	0.561	5.6	121	-0.01
4 S	2-Fluorophenol	0.949	1.077	-13.5	135	0.00
5 S	Phenol-d6	1.433	1.658	-15.7	140	0.00
6	bis(2-Chloroethyl)ether	1.385	1.343	3.0	114	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.190	0.254	-33.7#	190#	0.00
9	Naphthalene	4.250	4.094	3.7	124	0.00
10	Hexachlorobutadiene	0.166	0.156	6.0	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.565	2.6	128	0.00
12	2-Methylnaphthalene	0.664	0.649	2.3	129	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	135	0.00
14 S	2,4,6-Tribromophenol	0.044	0.065	-47.7#	226#	0.00
15 S	2-Fluorobiphenyl	1.587	1.421	10.5	126	-0.01
16	Acenaphthylene	1.655	1.507	8.9	142	0.00
17	Acenaphthene	1.140	1.068	6.3	138	0.00
18	Fluorene	1.471	1.317	10.5	136	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	-0.01
20	4-Bromophenyl-phenylether	0.196	0.184	6.1	126	-0.01
21	Hexachlorobenzene	0.213	0.205	3.8	125	0.00
22	Pentachlorophenol	0.015	0.019	-26.7#	831#	0.00
23	Phenanthrene	1.212	1.122	7.4	120	0.00
24	Anthracene	1.013	0.917	9.5	129	0.00
25 SURR	Fluoranthene-d10	4.728	3.785	19.9	108	0.00
26	Fluoranthene	1.446	1.158	19.9	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pvrene	1.300	1.343	-3.3	110	0.00
29 S	Terphenyl-d14	0.958	0.943	1.6	102	0.00
30	Benzo(a)anthracene	1.179	1.224	-3.8	116	-0.01
31	Chrysene	1.385	1.325	4.3	98	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.373	-28.4#	168#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.207	1.364	-13.0	127	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	123	-0.01
35	Benzo(b)fluoranthene	1.095	0.969	11.5	114	-0.01
36	Benzo(k)fluoranthene	1.228	1.049	14.6	110	-0.01
37 C	Benzo(a)pyrene	0.948	0.837	11.7	119	-0.01
38	Dibenzo(a,h)anthracene	0.943	0.844	10.5	123	-0.03
39	Benzo(a,h,i)perylene	1.009	0.927	8.1	120	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
Data File : BE100821.D
Acq On : 10 Dec 2019 22:02
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 05:43:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		