

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
 Data File : BE100823.D
 Acq On : 10 Dec 2019 23:17
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 05:41:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.729	0.635	12.9	104	0.00
3	n-Nitrosodimethylamine	0.594	0.530	10.8	123	0.00
4 S	2-Fluorophenol	0.949	1.094	-15.3	147	0.00
5 S	Phenol-d6	1.433	1.685	-17.6	152#	0.00
6	bis(2-Chloroethyl)ether	1.385	1.323	4.5	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
8 S	Nitrobenzene-d5	0.190	0.249	-31.1#	196#	0.00
9	Naphthalene	4.250	4.092	3.7	131	0.00
10	Hexachlorobutadiene	0.166	0.156	6.0	126	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.566	2.4	136	0.00
12	2-Methylnaphthalene	0.664	0.647	2.6	135	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	139	0.00
14 S	2,4,6-Tribromophenol	0.044	0.065	-47.7#	232#	0.00
15 S	2-Fluorobiphenyl	1.587	1.456	8.3	133	-0.01
16	Acenaphthylene	1.655	1.511	8.7	147	0.00
17	Acenaphthene	1.140	1.084	4.9	145	0.00
18	Fluorene	1.471	1.352	8.1	143	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	126	-0.01
20	4-Bromophenyl-phenylether	0.196	0.186	5.1	132	-0.01
21	Hexachlorobenzene	0.213	0.205	3.8	129	0.00
22	Pentachlorophenol	0.015	0.019	-26.7#	831#	0.00
23	Phenanthrene	1.212	1.150	5.1	127	0.00
24	Anthracene	1.013	0.907	10.5	131	0.00
25 SURR	Fluoranthene-d10	4.728	3.827	19.1	113	0.00
26	Fluoranthene	1.446	1.177	18.6	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
28	Pvrene	1.300	1.311	-0.8	113	0.00
29 S	Terphenyl-d14	0.958	0.935	2.4	107	0.00
30	Benzo(a)anthracene	1.179	1.169	0.8	117	-0.01
31	Chrysene	1.385	1.303	5.9	102	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.190	-11.3	154#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.207	1.341	-11.1	132	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	128	-0.01
35	Benzo(b)fluoranthene	1.095	0.933	14.8	115	-0.01
36	Benzo(k)fluoranthene	1.228	1.065	13.3	117	0.00
37 C	Benzo(a)pyrene	0.948	0.822	13.3	122	-0.01
38	Dibenzo(a,h)anthracene	0.943	0.864	8.4	132	-0.03
39	Benzo(a,h,i)perylene	1.009	0.911	9.7	123	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
Data File : BE100823.D
Acq On : 10 Dec 2019 23:17
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 05:41:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			