

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
 Data File : BE100823.D
 Acq On : 10 Dec 2019 23:17
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 05:41:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	120	0.00
2	1,4-Dioxane	0.400	0.348	13.0	104	0.00
3	n-Nitrosodimethylamine	0.400	0.357	10.8	123	0.00
4 S	2-Fluorophenol	0.400	0.461	-15.3	147	0.00
5 S	Phenol-d6	0.400	0.470	-17.5	152	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	131	0.00
8 S	Nitrobenzene-d5	0.400	0.523	-30.8#	196	0.00
9	Naphthalene	0.400	0.385	3.8	131	0.00
10	Hexachlorobutadiene	0.400	0.376	6.0	126	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	136	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	135	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	139	0.00
14 S	2,4,6-Tribromophenol	0.400	0.584	-46.0#	232	0.00
15 S	2-Fluorobiphenyl	0.400	0.367	8.3	133	-0.01
16	Acenaphthylene	0.400	0.365	8.8	147	0.00
17	Acenaphthene	0.400	0.380	5.0	145	0.00
18	Fluorene	0.400	0.368	8.0	143	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	126	-0.01
20	4-Bromophenyl-phenylether	0.400	0.380	5.0	132	-0.01
21	Hexachlorobenzene	0.400	0.385	3.8	129	0.00
22	Pentachlorophenol	0.400	0.510	-27.5#	831	0.00
23	Phenanthrene	0.400	0.380	5.0	127	0.00
24	Anthracene	0.400	0.358	10.5	131	0.00
25 SURR	Fluoranthene-d10	0.400	0.324	19.0	113	0.00
26	Fluoranthene	0.400	0.326	18.5	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
28	Pvrene	0.400	0.403	-0.8	113	0.00
29 S	Terphenyl-d14	0.400	0.390	2.5	107	0.00
30	Benzo(a)anthracene	0.400	0.397	0.8	117	-0.01
31	Chrysene	0.400	0.376	6.0	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.572	-43.0#	154	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.444	-11.0	132	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	128	-0.01
35	Benzo(b)fluoranthene	0.400	0.341	14.8	115	-0.01
36	Benzo(k)fluoranthene	0.400	0.347	13.3	117	0.00
37 C	Benzo(a)pyrene	0.400	0.347	13.3	122	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.366	8.5	132	-0.03
39	Benzo(a,h,i)perylene	0.400	0.361	9.8	123	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
Data File : BE100823.D
Acq On : 10 Dec 2019 23:17
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 05:41:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				