

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
 Data File : BE100839.D
 Acq On : 11 Dec 2019 9:08
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 10:33:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.729	0.720	1.2	109	0.00
3	n-Nitrosodimethylamine	0.594	0.586	1.3	125	0.00
4 S	2-Fluorophenol	0.949	1.068	-12.5	133	0.00
5 S	Phenol-d6	1.433	1.665	-16.2	139	0.00
6	bis(2-Chloroethyl)ether	1.385	1.346	2.8	114	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
8 S	Nitrobenzene-d5	0.190	0.248	-30.5#	184#	0.00
9	Naphthalene	4.250	4.113	3.2	124	0.00
10	Hexachlorobutadiene	0.166	0.154	7.2	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.580	0.568	2.1	128	0.00
12	2-Methylnaphthalene	0.664	0.643	3.2	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	148	0.00
14 S	2,4,6-Tribromophenol	0.044	0.119	-170.5#	455#	0.00
15 S	2-Fluorobiphenyl	1.587	1.295	18.4	127	-0.01
16	Acenaphthylene	1.655	1.507	8.9	157#	0.00
17	Acenaphthene	1.140	1.091	4.3	156#	0.00
18	Fluorene	1.471	1.761	-19.7	200#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	236#	-0.01
20	4-Bromophenyl-phenylether	0.196	0.169	13.8	224#	-0.01
21	Hexachlorobenzene	0.213	0.195	8.5	229#	0.00
22	Pentachlorophenol	0.015	0.025	-66.7#	2081#	0.00
23	Phenanthrene	1.212	1.137	6.2	235#	0.00
24	Anthracene	1.013	0.956	5.6	259#	0.00
25 SURR	Fluoranthene-d10	4.728	4.249	10.1	234#	0.00
26	Fluoranthene	1.446	1.318	8.9	238#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	218#	0.00
28	Pvrene	1.300	1.327	-2.1	238#	0.00
29 S	Terphenyl-d14	0.958	0.936	2.3	223#	0.00
30	Benzo(a)anthracene	1.179	1.189	-0.8	248#	-0.01
31	Chrysene	1.385	1.369	1.2	222#	0.00
32	Bis(2-ethylhexyl)phthalate	1.069	1.346	-25.9#	362#	0.00
33	Indeno(1,2,3-cd)pyrene	1.207	1.349	-11.8	276#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	263#	-0.01
35	Benzo(b)fluoranthene	1.095	0.938	14.3	238#	-0.01
36	Benzo(k)fluoranthene	1.228	1.086	11.6	245#	-0.01
37 C	Benzo(a)pyrene	0.948	0.835	11.9	255#	-0.01
38	Dibenzo(a,h)anthracene	0.943	0.893	5.3	281#	-0.03
39	Benzo(a,h,i)perylene	1.009	0.922	8.6	256#	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121019\
Data File : BE100839.D
Acq On : 11 Dec 2019 9:08
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 10:33:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			