

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE121117\
 Data File : BE094972.D
 Acq On : 11 Dec 2017 13:18
 Operator : SJ/JU
 Sample : I6740-07
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FT-PZ-456S-20171204

Manual Integrations
 APPROVED

Sohil
 12/11/2017 6:24:08 PM

Quant Time: Dec 11 15:52:48 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.48	152	7649	0.40	ng	0.00
7) Naphthalene-d8	11.33	136	17599	0.40	ng	0.00
13) Acenaphthene-d10	15.08	164	9819	0.40	ng	0.00
19) Phenanthrene-d10	17.79	188	24141	0.40	ng	0.01
27) Chrysene-d12	21.89	240	25117	0.40	ng	0.00
34) Perylene-d12	24.57	264	20175	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.95	112	2046	0.17	ng	0.00
5) Phenol-d6	7.60	99	2164	0.12	ng	0.01
8) Nitrobenzene-d5	9.67	82	6021	0.58	ng	0.00
11) 2-Methylnaphthalene-d10	12.87	152	10887	0.39	ng	0.00
14) 2,4,6-Tribromophenol	16.55	330	808	0.41	ng	0.02
15) 2-Fluorobiphenyl	13.71	172	27777	0.66	ng	0.00
25) Fluoranthene-d10	19.76	212	123892	0.38	ng	0.00
29) Terphenyl-d14	20.33	244	29180	0.63	ng	0.00

Target Compounds						Qvalue
22) Pentachlorophenol	17.43	266	272m	0.196	ng	
23) Phenanthrene	17.82	178	7374	0.096	ng	96
24) Anthracene	17.91	178	2163	0.027	ng	# 71
26) Fluoranthene	19.79	202	5699	0.058	ng	99
28) Pyrene	20.14	202	4771	0.055	ng	95
30) Benzo(a)anthracene	21.87	228	5332	0.069	ng	# 64
31) Chrysene	21.92	228	6724	0.079	ng	# 66
32) Bis(2-ethylhexyl)phthalate	21.75	149	5251	0.041	ng	# 100
33) Indeno(1,2,3-cd)pyrene	27.44	276	4913	0.071	ng	99
35) Benzo(b)fluoranthene	23.73	252	5504	0.075	ng	# 49
36) Benzo(k)fluoranthene	23.79	252	4862	0.065	ng	# 48
37) Benzo(a)pyrene	24.44	252	4287	0.066	ng	# 51
38) Dibenzo(a,h)anthracene	27.47	278	3944	0.066	ng	# 52
39) Benzo(g,h,i)perylene	28.32	276	4294	0.072	ng	# 59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE121117\
Data File : BE094972.D
Acq On : 11 Dec 2017 13:18
Operator : SJ/JU
Sample : I6740-07
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FT-PZ-456S-20171204

Manual Integrations
APPROVED

Sohil
12/11/2017 6:24:08 PM

Quant Time: Dec 11 15:52:48 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

