

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE121117\
 Data File : BE094974.D
 Acq On : 11 Dec 2017 15:26
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 16:16:58 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.419	0.470	-12.2	97	-0.01
3	n-Nitrosodimethylamine	0.533	0.474	11.1	83	-0.01
4 S	2-Fluorophenol	0.633	0.592	6.5	85	0.00
5 S	Phenol-d6	0.938	0.916	2.3	90	0.01
6	bis(2-Chloroethyl)ether	0.930	0.927	0.3	90	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.237	0.346	-46.0#	138	0.00
9	Naphthalene	4.690	4.716	-0.6	94	0.00
10	Hexachlorobutadiene	0.198	0.211	-6.6	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.652	-2.2	95	0.00
12	2-Methylnaphthalene	1.165	1.168	-0.3	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.086	0.085	1.2	101	0.01
15 S	2-Fluorobiphenyl	1.713	1.716	-0.2	97	0.00
16	Acenaphthylene	2.260	2.034	10.0	91	0.00
17	Acenaphthene	1.461	1.389	4.9	95	0.00
18	Fluorene	1.859	1.705	8.3	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.01
20	4-Bromophenyl-phenylether	0.223	0.223	0.0	93	0.01
21	Hexachlorobenzene	0.218	0.226	-3.7	94	0.01
22	Pentachlorophenol	0.054	0.047	13.0	88	0.00
23	Phenanthrene	1.275	1.243	2.5	88	0.00
24	Anthracene	1.351	1.251	7.4	87	0.00
25 SURR	Fluoranthene-d10	5.466	4.994	8.6	84	0.00
26	Fluoranthene	1.638	1.496	8.7	84	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
28	Pvrene	1.381	1.346	2.5	82	0.00
29 S	Terphenyl-d14	0.739	0.739	0.0	85	0.00
30	Benzo(a)anthracene	1.227	1.115	9.1	78	0.00
31	Chrysene	1.355	1.305	3.7	81	0.00
32	Bis(2-ethylhexyl)phthalate	2.033	1.442	29.1#	69	-0.02
33	Indeno(1,2,3-cd)pyrene	1.109	1.058	4.6	81	0.00
34 I	Perylene-d12	1.000	1.000	0.0	83	0.00
35	Benzo(b)fluoranthene	1.460	1.432	1.9	83	0.00
36	Benzo(k)fluoranthene	1.489	1.377	7.5	79	0.00
37 C	Benzo(a)pyrene	1.280	1.246	2.7	83	0.00
38	Dibenzo(a,h)anthracene	1.183	1.142	3.5	82	0.00
39	Benzo(a,h,i)perylene	1.181	1.138	3.6	81	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE121117\
Data File : BE094974.D
Acq On : 11 Dec 2017 15:26
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 16:16:58 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			