

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE121117\
 Data File : BE094974.D
 Acq On : 11 Dec 2017 15:26
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 16:16:58 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 13:33:28 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	92	0.00
2	1,4-Dioxane	0.400	0.449	-12.2	97	-0.01
3	n-Nitrosodimethylamine	0.400	0.355	11.3	83	-0.01
4 S	2-Fluorophenol	0.400	0.374	6.5	85	0.00
5 S	Phenol-d6	0.400	0.391	2.3	90	0.01
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	90	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	94	0.00
8 S	Nitrobenzene-d5	0.400	0.585	-46.2#	138	0.00
9	Naphthalene	0.400	0.402	-0.5	94	0.00
10	Hexachlorobutadiene	0.400	0.426	-6.5	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.408	-2.0	95	0.00
12	2-Methylnaphthalene	0.400	0.401	-0.3	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.400	0.420	-5.0	101	0.01
15 S	2-Fluorobiphenyl	0.400	0.401	-0.3	97	0.00
16	Acenaphthylene	0.400	0.360	10.0	91	0.00
17	Acenaphthene	0.400	0.380	5.0	95	0.00
18	Fluorene	0.400	0.367	8.3	89	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	91	0.01
20	4-Bromophenyl-phenylether	0.400	0.400	0.0	93	0.01
21	Hexachlorobenzene	0.400	0.415	-3.7	94	0.01
22	Pentachlorophenol	0.400	0.393	1.8	88	0.00
23	Phenanthrene	0.400	0.390	2.5	88	0.00
24	Anthracene	0.400	0.370	7.5	87	0.00
25 SURR	Fluoranthene-d10	0.400	0.365	8.8	84	0.00
26	Fluoranthene	0.400	0.365	8.8	84	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	84	0.00
28	Pvrene	0.400	0.390	2.5	82	0.00
29 S	Terphenyl-d14	0.400	0.400	0.0	85	0.00
30	Benzo(a)anthracene	0.400	0.364	9.0	78	0.00
31	Chrysene	0.400	0.385	3.8	81	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.284	29.0#	69	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.382	4.5	81	0.00
34 I	Perylene-d12	0.400	0.400	0.0	83	0.00
35	Benzo(b)fluoranthene	0.400	0.392	2.0	83	0.00
36	Benzo(k)fluoranthene	0.400	0.370	7.5	79	0.00
37 C	Benzo(a)pyrene	0.400	0.389	2.8	83	0.00
38	Dibenzo(a,h)anthracene	0.400	0.386	3.5	82	0.00
39	Benzo(a,h,i)perylene	0.400	0.385	3.8	81	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE121117\
Data File : BE094974.D
Acq On : 11 Dec 2017 15:26
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 11 16:16:58 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 13:33:28 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			