

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121118\
 Data File : BE098334.D
 Acq On : 11 Dec 2018 15:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 12 01:37:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.692	0.640	7.5	105	0.00
3	n-Nitrosodimethylamine	0.623	0.427	31.5#	82	0.00
4 S	2-Fluorophenol	0.936	0.870	7.1	107	0.00
5 S	Phenol-d6	1.326	1.218	8.1	101	0.00
6	bis(2-Chloroethyl)ether	1.260	1.168	7.3	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.364	0.348	4.4	112	0.01
9	Naphthalene	4.025	3.923	2.5	108	0.00
10	Hexachlorobutadiene	0.256	0.251	2.0	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.608	6.5	102	0.00
12	2-Methylnaphthalene	0.654	0.611	6.6	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.147	0.126	14.3	87	0.00
15 S	2-Fluorobiphenyl	1.687	1.676	0.7	104	0.00
16	Acenaphthylene	1.874	1.863	0.6	98	0.00
17	Acenaphthene	1.126	1.118	0.7	96	0.00
18	Fluorene	1.506	1.420	5.7	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.215	0.202	6.0	86	0.00
21	Hexachlorobenzene	0.231	0.232	-0.4	90	0.00
22	Pentachlorophenol	0.046	0.031	32.6#	76	0.00
23	Phenanthrene	0.972	0.936	3.7	86	0.00
24	Anthracene	0.866	0.772	10.9	79	0.00
25 SURR	Fluoranthene-d10	5.131	4.912	4.3	81	0.00
26	Fluoranthene	1.286	1.231	4.3	81	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	1.255	1.090	13.1	82	0.00
29 S	Terphenyl-d14	0.972	0.799	17.8	80	0.00
30	Benzo(a)anthracene	1.139	1.042	8.5	90	0.00
31	Chrysene	1.193	1.159	2.8	95	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.950	15.7	82	0.00
33	Indeno(1,2,3-cd)pyrene	1.187	1.215	-2.4	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.094	1.064	2.7	104	0.00
36	Benzo(k)fluoranthene	1.274	1.322	-3.8	112	0.00
37 C	Benzo(a)pyrene	1.129	1.121	0.7	105	0.00
38	Dibenzo(a,h)anthracene	1.063	1.002	5.7	100	0.00
39	Benzo(a,h,i)perylene	1.071	1.038	3.1	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121118\
Data File : BE098334.D
Acq On : 11 Dec 2018 15:50
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 12 01:37:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			