

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121118\
 Data File : BE098352.D
 Acq On : 12 Dec 2018 3:33
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 12 04:09:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.692	0.712	-2.9	123	0.00
3	n-Nitrosodimethylamine	0.623	0.466	25.2#	95	0.00
4 S	2-Fluorophenol	0.936	0.934	0.2	121	0.00
5 S	Phenol-d6	1.326	1.249	5.8	110	0.00
6	bis(2-Chloroethyl)ether	1.260	1.180	6.3	114	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	-0.01
8 S	Nitrobenzene-d5	0.364	0.377	-3.6	127	0.00
9	Naphthalene	4.025	4.014	0.3	115	0.00
10	Hexachlorobutadiene	0.256	0.254	0.8	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.622	4.3	109	0.00
12	2-Methylnaphthalene	0.654	0.612	6.4	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.147	0.153	-4.1	120	0.00
15 S	2-Fluorobiphenyl	1.687	1.577	6.5	112	0.00
16	Acenaphthylene	1.874	1.838	1.9	111	0.00
17	Acenaphthene	1.126	1.092	3.0	107	0.00
18	Fluorene	1.506	1.487	1.3	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	-0.01
20	4-Bromophenyl-phenylether	0.215	0.199	7.4	100	0.00
21	Hexachlorobenzene	0.231	0.224	3.0	103	0.00
22	Pentachlorophenol	0.046	0.035	23.9	100	0.00
23	Phenanthrene	0.972	0.950	2.3	103	0.00
24	Anthracene	0.866	0.777	10.3	95	0.00
25 SURR	Fluoranthene-d10	5.131	5.083	0.9	100	0.00
26	Fluoranthene	1.286	1.288	-0.2	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
28	Pvrene	1.255	1.057	15.8	102	0.00
29 S	Terphenyl-d14	0.972	0.798	17.9	102	0.00
30	Benzo(a)anthracene	1.139	1.050	7.8	116	-0.01
31	Chrysene	1.193	1.130	5.3	118	0.00
32	Bis(2-ethylhexyl)phthalate	2.313	1.978	14.5	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.187	1.166	1.8	128	0.00
34 I	Perylene-d12	1.000	1.000	0.0	130	0.00
35	Benzo(b)fluoranthene	1.094	1.053	3.7	129	0.00
36	Benzo(k)fluoranthene	1.274	1.305	-2.4	139	0.00
37 C	Benzo(a)pyrene	1.129	1.092	3.3	128	0.00
38	Dibenzo(a,h)anthracene	1.063	1.021	4.0	129	0.00
39	Benzo(a,h,i)perylene	1.071	1.014	5.3	127	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121118\
Data File : BE098352.D
Acq On : 12 Dec 2018 3:33
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 12 04:09:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			