

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
 Data File : BE100848.D
 Acq On : 11 Dec 2019 16:16
 Operator : JU
 Sample : K6185-14
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 TT101D-20191205

Quant Time: Dec 12 04:55:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 17:18:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.76	152	5199	0.40	ng	-0.01
7) Naphthalene-d8	10.54	136	23573	0.40	ng	0.00
13) Acenaphthene-d10	14.40	164	12996	0.40	ng	0.00
19) Phenanthrene-d10	17.12	188	30987	0.40	ng	-0.01
27) Chrysene-d12	21.31	240	25932	0.40	ng	0.00
34) Perylene-d12	23.75	264	27170	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.37	112	2535	0.21	ng	0.00
5) Phenol-d6	6.93	99	2414	0.13	ng	0.00
8) Nitrobenzene-d5	8.90	82	6152	0.55	ng	0.00
11) 2-Methylnaphthalene-d10	12.14	152	15303	0.45	ng	0.00
14) 2,4,6-Tribromophenol	15.86	330	863	0.60	ng	-0.01
15) 2-Fluorobiphenyl	13.02	172	26610	0.52	ng	-0.01
25) Fluoranthene-d10	19.16	212	149914	0.41	ng	0.00
29) Terphenyl-d14	19.76	244	31885	0.51	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.34	88	52124	5.499	ng	Qvalue # 89
6) bis(2-Chloroethyl)ether	7.18	93	1222	0.068	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE121119\
Data File : BE100848.D
Acq On : 11 Dec 2019 16:16
Operator : JU
Sample : K6185-14
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
TT101D-20191205

Quant Time: Dec 12 04:55:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 17:18:11 2019
Response via : Initial Calibration

